

Data Path : Z:\voasrv\HPCHEM1\MSVOA_D\Data\VD101121\
 Data File : VD070658.D
 Acq On : 11 Oct 2021 14:16
 Operator : VA/SY
 Sample : M4136-05
 Misc : 5.18G/5.00ml/MSVOA_D/SOIL
 ALS Vial : 8 Sample Multiplier: 1

Instrument :
 MSVOA_D
 ClientSampleId :
 PSC124042

Quant Time: Oct 12 05:40:27 2021
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_D\Method\82D100521S.M
 Quant Title : SW846 8260
 QLast Update : Wed Oct 06 09:10:31 2021
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) Pentafluorobenzene	7.979	168	134256	50.00	ug/l	0.00
34) 1,4-Difluorobenzene	8.861	114	252581	50.00	ug/l	0.00
63) Chlorobenzene-d5	11.638	117	280934	50.00	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	13.561	152	127227	50.00	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	8.332	65	112238	71.04	ug/l	0.00
Spiked Amount	50.000	Range	50 - 163	Recovery	=	142.08%
35) Dibromofluoromethane	7.914	113	94252	56.13	ug/l	0.00
Spiked Amount	50.000	Range	54 - 147	Recovery	=	112.26%
50) Toluene-d8	10.338	98	333668	51.98	ug/l	0.00
Spiked Amount	50.000	Range	49 - 140	Recovery	=	103.96%
62) 4-Bromofluorobenzene	12.626	95	129899	60.54	ug/l	0.00
Spiked Amount	50.000	Range	25 - 144	Recovery	=	121.08%

Target Compounds	Qvalue

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_D\Data\VD101121\
Data File : VD070658.D
Acq On : 11 Oct 2021 14:16
Operator : VA/SY
Sample : M4136-05
Misc : 5.18G/5.00ml/MSVOA_D/SOIL
ALS Vial : 8 Sample Multiplier: 1

Instrument :
MSVOA_D
ClientSampleId :
PSC124042

Quant Time: Oct 12 05:40:27 2021
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_D\Method\82D100521S.M
Quant Title : SW846 8260
QLast Update : Wed Oct 06 09:10:31 2021
Response via : Initial Calibration

