

Data Path : Z:\voasrv\HPCHEM1\MSVOA_D\Data\VD101121\
 Data File : VD070659.D
 Acq On : 11 Oct 2021 14:44
 Operator : VA/SY
 Sample : M4135-01
 Misc : 6.27G/5.00ml/MSVOA_D/SOIL
 ALS Vial : 9 Sample Multiplier: 1

Instrument :
 MSVOA_D
 ClientSampleId :
 TS-2

Quant Time: Oct 12 05:40:45 2021
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_D\Method\82D100521S.M
 Quant Title : SW846 8260
 QLast Update : Wed Oct 06 09:10:31 2021
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) Pentafluorobenzene	7.979	168	376819	50.00	ug/l	0.00
34) 1,4-Difluorobenzene	8.861	114	731201	50.00	ug/l	0.00
63) Chlorobenzene-d5	11.638	117	650829	50.00	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	13.561	152	220575	50.00	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	8.326	65	249280	56.21	ug/l	0.00
Spiked Amount 50.000	Range 50 - 163		Recovery = 112.42%			
35) Dibromofluoromethane	7.914	113	243645	50.12	ug/l	0.00
Spiked Amount 50.000	Range 54 - 147		Recovery = 100.24%			
50) Toluene-d8	10.338	98	909460	48.94	ug/l	0.00
Spiked Amount 50.000	Range 49 - 140		Recovery = 97.88%			
62) 4-Bromofluorobenzene	12.626	95	264667	42.61	ug/l	0.00
Spiked Amount 50.000	Range 25 - 144		Recovery = 85.22%			
Target Compounds						Qvalue
30) Chloroform	7.709	83	11977	1.24	ug/l	94

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_D\Data\VD101121\
Data File : VD070659.D
Acq On : 11 Oct 2021 14:44
Operator : VA/SY
Sample : M4135-01
Misc : 6.27G/5.00ml/MSVOA_D/SOIL
ALS Vial : 9 Sample Multiplier: 1

Instrument :
MSVOA_D
ClientSampleId :
TS-2

Quant Time: Oct 12 05:40:45 2021
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_D\Method\82D100521S.M
Quant Title : SW846 8260
QLast Update : Wed Oct 06 09:10:31 2021
Response via : Initial Calibration

