

Data Path : Z:\voasrv\HPCHEM1\MSVOA_D\Data\VD101121\
 Data File : VD070661.D
 Acq On : 11 Oct 2021 15:41
 Operator : VA/SY
 Sample : M4117-05
 Misc : 5.44G/5.00ml/MSVOA_D/SOIL
 ALS Vial : 11 Sample Multiplier: 1

Instrument :
 MSVOA_D
 ClientSampleId :
 SB-2(1.5-2.0)

Quant Time: Oct 12 05:41:19 2021
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_D\Method\82D100521S.M
 Quant Title : SW846 8260
 QLast Update : Wed Oct 06 09:10:31 2021
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) Pentafluorobenzene	7.979	168	391631	50.00	ug/l	0.00
34) 1,4-Difluorobenzene	8.861	114	773324	50.00	ug/l	0.00
63) Chlorobenzene-d5	11.638	117	711725	50.00	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	13.561	152	303950	50.00	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	8.326	65	247557	53.71	ug/l	0.00
Spiked Amount	50.000	Range	50 - 163	Recovery	=	107.42%
35) Dibromofluoromethane	7.908	113	251082	48.83	ug/l	0.00
Spiked Amount	50.000	Range	54 - 147	Recovery	=	97.66%
50) Toluene-d8	10.332	98	989127	50.32	ug/l	0.00
Spiked Amount	50.000	Range	49 - 140	Recovery	=	100.64%
62) 4-Bromofluorobenzene	12.626	95	310479	47.26	ug/l	0.00
Spiked Amount	50.000	Range	25 - 144	Recovery	=	94.52%

Target Compounds	Qvalue

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_D\Data\VD101121\
Data File : VD070661.D
Acq On : 11 Oct 2021 15:41
Operator : VA/SY
Sample : M4117-05
Misc : 5.44G/5.00ml/MSVOA_D/SOIL
ALS Vial : 11 Sample Multiplier: 1

Instrument :
MSVOA_D
ClientSampleId :
SB-2(1.5-2.0)

Quant Time: Oct 12 05:41:19 2021
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_D\Method\82D100521S.M
Quant Title : SW846 8260
QLast Update : Wed Oct 06 09:10:31 2021
Response via : Initial Calibration

