

Data Path : Z:\voasrv\HPCHEM1\MSVOA_D\Data\VD101121\
 Data File : VD070665.D
 Acq On : 11 Oct 2021 17:35
 Operator : VA/SY
 Sample : M4164-01
 Misc : 4.96G/5.00ml/MSVOA_D/SOIL
 ALS Vial : 15 Sample Multiplier: 1

Instrument :
 MSVOA_D
 ClientSampleId :
 BU-03-100821

Quant Time: Oct 12 05:42:31 2021
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_D\Method\82D100521S.M
 Quant Title : SW846 8260
 QLast Update : Wed Oct 06 09:10:31 2021
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) Pentafluorobenzene	7.979	168	360759	50.00	ug/l	0.00
34) 1,4-Difluorobenzene	8.861	114	709393	50.00	ug/l	0.00
63) Chlorobenzene-d5	11.638	117	658712	50.00	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	13.561	152	253240	50.00	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	8.326	65	209743	49.40	ug/l	0.00
Spiked Amount	50.000	Range	50 - 163	Recovery	=	98.80%
35) Dibromofluoromethane	7.909	113	228169	48.38	ug/l	0.00
Spiked Amount	50.000	Range	54 - 147	Recovery	=	96.76%
50) Toluene-d8	10.332	98	904318	50.16	ug/l	0.00
Spiked Amount	50.000	Range	49 - 140	Recovery	=	100.32%
62) 4-Bromofluorobenzene	12.620	95	260882	43.29	ug/l	0.00
Spiked Amount	50.000	Range	25 - 144	Recovery	=	86.58%

Target Compounds	Qvalue

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_D\Data\VD101121\
Data File : VD070665.D
Acq On : 11 Oct 2021 17:35
Operator : VA/SY
Sample : M4164-01
Misc : 4.96G/5.00ml/MSVOA_D/SOIL
ALS Vial : 15 Sample Multiplier: 1

Instrument :
MSVOA_D
ClientSampleId :
BU-03-100821

Quant Time: Oct 12 05:42:31 2021
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_D\Method\82D100521S.M
Quant Title : SW846 8260
QLast Update : Wed Oct 06 09:10:31 2021
Response via : Initial Calibration

