

Data Path : Z:\voasrv\HPCHEM1\MSVOA_D\Data\VD101223\
 Data File : VD077320.D
 Acq On : 12 Oct 2023 16:42
 Operator : JC/SY
 Sample : 04837-05RE
 Misc : 5.02G/5.0ml/MSVOA_D/SOIL
 ALS Vial : 8 Sample Multiplier: 1

Instrument :
 MSVOA_D
 ClientSampleId :
 ETGI-290RE

Quant Time: Oct 13 02:32:04 2023
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_D\Method\82D101123S.M
 Quant Title : SW846 8260
 QLast Update : Thu Oct 12 14:39:07 2023
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) Pentafluorobenzene	7.875	168	189602	50.000	ug/l	# 0.00
34) 1,4-Difluorobenzene	8.775	114	442114	50.000	ug/l	0.00
63) Chlorobenzene-d5	11.581	117	569283	50.000	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	13.522	152	275174	50.000	ug/l	0.00

System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	8.228	65	95826	64.286	ug/l	0.00
Spiked Amount	50.000	Range	50 - 163	Recovery	=	128.580%
35) Dibromofluoromethane	7.810	113	133424	57.955	ug/l	0.00
Spiked Amount	50.000	Range	54 - 147	Recovery	=	115.900%
50) Toluene-d8	10.269	98	573515	74.292	ug/l	0.00
Spiked Amount	50.000	Range	58 - 134	Recovery	=	148.580%#
62) 4-Bromofluorobenzene	12.575	95	244749	94.396	ug/l	0.00
Spiked Amount	50.000	Range	30 - 143	Recovery	=	188.800%#

Target Compounds	Qvalue

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_D\Data\VD101223\
Data File : VD077320.D
Acq On : 12 Oct 2023 16:42
Operator : JC/SY
Sample : 04837-05RE
Misc : 5.02G/5.0ml/MSVOA_D/SOIL
ALS Vial : 8 Sample Multiplier: 1

Instrument :
MSVOA_D
ClientSampleId :
ETGI-290RE

Quant Time: Oct 13 02:32:04 2023
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_D\Method\82D101123S.M
Quant Title : SW846 8260
QLast Update : Thu Oct 12 14:39:07 2023
Response via : Initial Calibration

