

Data Path : Z:\voasrv\HPCHEM1\MSVOA_D\Data\VD101223\
 Data File : VD077322.D
 Acq On : 12 Oct 2023 17:37
 Operator : JC/SY
 Sample : 04830-01RE
 Misc : 5.06G/5.0ml/MSVOA_D/SOIL
 ALS Vial : 10 Sample Multiplier: 1

Instrument :
 MSVOA_D
 ClientSampleId :
 RT-2327RE

Quant Time: Oct 13 02:32:43 2023
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_D\Method\82D101123S.M
 Quant Title : SW846 8260
 QLast Update : Thu Oct 12 14:39:07 2023
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) Pentafluorobenzene	7.869	168	10781	50.000	ug/l	# 0.00
34) 1,4-Difluorobenzene	8.781	114	21890	50.000	ug/l	0.00
63) Chlorobenzene-d5	11.581	117	31493	50.000	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	13.516	152	16354	50.000	ug/l	0.00

System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	8.222	65	7086	83.603	ug/l	0.00
Spiked Amount	50.000	Range	50 - 163	Recovery	=	167.200%#
35) Dibromofluoromethane	7.810	113	8532	74.850	ug/l	0.00
Spiked Amount	50.000	Range	54 - 147	Recovery	=	149.700%#
50) Toluene-d8	10.263	98	30097	78.743	ug/l	0.00
Spiked Amount	50.000	Range	58 - 134	Recovery	=	157.480%#
62) 4-Bromofluorobenzene	12.575	95	13387	104.281	ug/l	0.00
Spiked Amount	50.000	Range	30 - 143	Recovery	=	208.560%#

Target Compounds	Qvalue

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_D\Data\VD101223\
Data File : VD077322.D
Acq On : 12 Oct 2023 17:37
Operator : JC/SY
Sample : 04830-01RE
Misc : 5.06G/5.0ml/MSVOA_D/SOIL
ALS Vial : 10 Sample Multiplier: 1

Instrument :
MSVOA_D
ClientSampleId :
RT-2327RE

Quant Time: Oct 13 02:32:43 2023
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_D\Method\82D101123S.M
Quant Title : SW846 8260
QLast Update : Thu Oct 12 14:39:07 2023
Response via : Initial Calibration

