


Data Path : Z:\voasrv\HPCHEM1\MSVOA D\Data\VD101320\
 Data File : VD067242.D
 Acq On : 13 Oct 2020 09:27
 Operator : VA/SY
 Sample : BFB
 Misc : 5.00G/5.00ml/MSVOA D/SOIL
 ALS Vial : 1 Sample Multiplier: 1

Instrument :
 MSVOA_D
 ClientSampleId :
 BFB

Integration File: RTEINT.P

Method : Z:\VOASRV\HPCHEM1\MSVOA_D\METHOD\82D101320S.M
 Title : SW846 8260
 Last Update : Wed Oct 14 03:30:30 2020



AutoFind: Scans 1821, 1822, 1823; Background Corrected with Scan 1813

Target Mass	Rel. to Mass	Lower Limit%	Upper Limit%	Rel. Abn%	Raw Abn	Result Pass/Fail
50	95	15	40	17.0	34573	PASS
75	95	30	60	51.2	103840	PASS
95	95	100	100	100.0	202944	PASS
96	95	5	9	6.1	12373	PASS
173	174	0.00	2	0.0	0	PASS
174	95	50	100	86.3	175061	PASS
175	174	5	9	8.0	14018	PASS
176	174	95	101	96.2	168469	PASS
177	176	5	9	7.4	12490	PASS