

Data Path : Z:\voasrv\HPCHEM1\MSVOA_D\Data\VD101321\
 Data File : VD070693.D
 Acq On : 13 Oct 2021 13:04
 Operator : VA/SY
 Sample : VD1013SBL01
 Misc : 5.00G/5.00ml/MSVOA_D/SOIL
 ALS Vial : 3 Sample Multiplier: 1

Instrument :
 MSVOA_D
 ClientSampleId :
 VD1013SBL01

Quant Time: Oct 14 04:02:27 2021
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_D\Method\82D100521S.M
 Quant Title : SW846 8260
 QLast Update : Wed Oct 06 09:10:31 2021
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) Pentafluorobenzene	7.979	168	452717	50.000	ug/l	0.00
34) 1,4-Difluorobenzene	8.861	114	851086	50.000	ug/l	0.00
63) Chlorobenzene-d5	11.637	117	791365	50.000	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	13.561	152	334494	50.000	ug/l	0.00

System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	8.326	65	256999	48.237	ug/l	0.00
Spiked Amount	50.000	Range	50 - 163	Recovery	=	96.480%
35) Dibromofluoromethane	7.914	113	274118	48.443	ug/l	0.00
Spiked Amount	50.000	Range	54 - 147	Recovery	=	96.880%
50) Toluene-d8	10.337	98	1076254	49.754	ug/l	0.00
Spiked Amount	50.000	Range	49 - 140	Recovery	=	99.500%
62) 4-Bromofluorobenzene	12.620	95	331354	45.831	ug/l	0.00
Spiked Amount	50.000	Range	25 - 144	Recovery	=	91.660%

Target Compounds	Qvalue

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_D\Data\VD101321\
Data File : VD070693.D
Acq On : 13 Oct 2021 13:04
Operator : VA/SY
Sample : VD1013SBL01
Misc : 5.00G/5.00ml/MSVOA_D/SOIL
ALS Vial : 3 Sample Multiplier: 1

Instrument :
MSVOA_D
ClientSampleId :
VD1013SBL01

Quant Time: Oct 14 04:02:27 2021
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_D\Method\82D100521S.M
Quant Title : SW846 8260
QLast Update : Wed Oct 06 09:10:31 2021
Response via : Initial Calibration

