

Data Path : Z:\voasrv\HPCHEM1\MSVOA_D\Data\VD101321\
 Data File : VD070696.D
 Acq On : 13 Oct 2021 14:28
 Operator : VA/SY
 Sample : M4191-01
 Misc : 5.11G/5.00ml/MSVOA_D/SOIL
 ALS Vial : 6 Sample Multiplier: 1

Instrument :
 MSVOA_D
 ClientSampleId :
 COMP-4

Quant Time: Oct 14 04:03:56 2021
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_D\Method\82D100521S.M
 Quant Title : SW846 8260
 QLast Update : Wed Oct 06 09:10:31 2021
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) Pentafluorobenzene	7.979	168	324030	50.000	ug/l	0.00
34) 1,4-Difluorobenzene	8.861	114	621774	50.000	ug/l	0.00
63) Chlorobenzene-d5	11.638	117	575762	50.000	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	13.561	152	239245	50.000	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	8.326	65	203543	53.376	ug/l	0.00
Spiked Amount	50.000	Range	50 - 163	Recovery	=	106.760%
35) Dibromofluoromethane	7.914	113	205056	49.603	ug/l	0.00
Spiked Amount	50.000	Range	54 - 147	Recovery	=	99.200%
50) Toluene-d8	10.338	98	779150	49.303	ug/l	0.00
Spiked Amount	50.000	Range	49 - 140	Recovery	=	98.600%
62) 4-Bromofluorobenzene	12.626	95	243409	46.084	ug/l	0.00
Spiked Amount	50.000	Range	25 - 144	Recovery	=	92.160%
Target Compounds						
16) Acetone	4.162	43	8591	13.267	ug/l	Qvalue 96

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_D\Data\VD101321\
Data File : VD070696.D
Acq On : 13 Oct 2021 14:28
Operator : VA/SY
Sample : M4191-01
Misc : 5.11G/5.00ml/MSVOA_D/SOIL
ALS Vial : 6 Sample Multiplier: 1

Instrument :
MSVOA_D
ClientSampleId :
COMP-4

Quant Time: Oct 14 04:03:56 2021
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_D\Method\82D100521S.M
Quant Title : SW846 8260
QLast Update : Wed Oct 06 09:10:31 2021
Response via : Initial Calibration

