

Data Path : Z:\voasrv\HPCHEM1\MSVOA_D\Data\VD101321\
 Data File : VD070704.D
 Acq On : 13 Oct 2021 18:14
 Operator : VA/SY
 Sample : M4194-04
 Misc : 5.72G/5.00ml/MSVOA_D/SOIL
 ALS Vial : 14 Sample Multiplier: 1

Instrument :
 MSVOA_D
 ClientSampleId :
 VNJ-209

Quant Time: Oct 14 04:05:19 2021
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_D\Method\82D100521S.M
 Quant Title : SW846 8260
 QLast Update : Wed Oct 06 09:10:31 2021
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) Pentafluorobenzene	7.973	168	346253	50.000	ug/l	0.00
34) 1,4-Difluorobenzene	8.861	114	660209	50.000	ug/l	0.00
63) Chlorobenzene-d5	11.637	117	623585	50.000	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	13.561	152	258144	50.000	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	8.326	65	220568	54.128	ug/l	0.00
Spiked Amount	50.000	Range	50 - 163	Recovery	=	108.260%
35) Dibromofluoromethane	7.908	113	217888	49.639	ug/l	0.00
Spiked Amount	50.000	Range	54 - 147	Recovery	=	99.280%
50) Toluene-d8	10.332	98	829901	49.457	ug/l	0.00
Spiked Amount	50.000	Range	49 - 140	Recovery	=	98.920%
62) 4-Bromofluorobenzene	12.620	95	259163	46.210	ug/l	0.00
Spiked Amount	50.000	Range	25 - 144	Recovery	=	92.420%

Target Compounds	Qvalue

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_D\Data\VD101321\
Data File : VD070704.D
Acq On : 13 Oct 2021 18:14
Operator : VA/SY
Sample : M4194-04
Misc : 5.72G/5.00ml/MSVOA_D/SOIL
ALS Vial : 14 Sample Multiplier: 1

Instrument :
MSVOA_D
ClientSampleId :
VNJ-209

Quant Time: Oct 14 04:05:19 2021
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_D\Method\82D100521S.M
Quant Title : SW846 8260
QLast Update : Wed Oct 06 09:10:31 2021
Response via : Initial Calibration

