


Data Path : Z:\voasrv\HPCHEM1\MSVOA D\Data\VD101519\
 Data File : VD063962.D
 Acq On : 15 Oct 2019 09:34
 Operator : VA/SY
 Sample : BFB
 Misc : 5.00G/5.00ml/MSVOA D/SOIL
 ALS Vial : 1 Sample Multiplier: 1

Instrument :
 MSVOA_D
 ClientSampleId :
 BFB

Integration File: RTEINT.P

Method : Z:\VOASRV\HPCHEM1\MSVOA_D\METHOD\82D100219S.M
 Title : SW846 8260
 Last Update : Sat Oct 05 03:51:17 2019



AutoFind: Scans 1821, 1822, 1823; Background Corrected with Scan 1813

Target Mass	Rel. to Mass	Lower Limit%	Upper Limit%	Rel. Abn%	Raw Abn	Result Pass/Fail
50	95	15	40	20.2	19472	PASS
75	95	30	60	55.6	53602	PASS
95	95	100	100	100.0	96424	PASS
96	95	5	9	6.7	6490	PASS
173	174	0.00	2	2.0	1833	PASS
174	95	50	100	96.6	93189	PASS
175	174	5	9	7.9	7374	PASS
176	174	95	101	98.3	91648	PASS
177	176	5	9	6.9	6345	PASS