

Data Path : Z:\voasrv\HPCHEM1\MSVOA_D\Data\VD101521\
 Data File : VD070717.D
 Acq On : 15 Oct 2021 15:17
 Operator : VA/SY
 Sample : VD1015SBL01
 Misc : 5.00G/5.00ml/MSVOA_D/SOIL
 ALS Vial : 12 Sample Multiplier: 1

Instrument :
 MSVOA_D
 ClientSampleId :
 VD1015SBL01

Quant Time: Oct 18 08:32:59 2021
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_D\Method\82D101521S.M
 Quant Title : SW846 8260
 QLast Update : Mon Oct 18 08:32:03 2021
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) Pentafluorobenzene	7.973	168	506501	50.00	ug/l	0.00
34) 1,4-Difluorobenzene	8.861	114	908229	50.00	ug/l	0.00
63) Chlorobenzene-d5	11.638	117	845504	50.00	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	13.561	152	354111	50.00	ug/l	0.00

System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	8.326	65	318792	53.14	ug/l	0.00
Spiked Amount	50.000	Range	50 - 163	Recovery	=	106.28%
35) Dibromofluoromethane	7.908	113	300961	51.75	ug/l	0.00
Spiked Amount	50.000	Range	54 - 147	Recovery	=	103.50%
50) Toluene-d8	10.332	98	1165936	52.30	ug/l	0.00
Spiked Amount	50.000	Range	49 - 140	Recovery	=	104.60%
62) 4-Bromofluorobenzene	12.620	95	371796	48.44	ug/l	0.00
Spiked Amount	50.000	Range	25 - 144	Recovery	=	96.88%

Target Compounds	Qvalue

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_D\Data\VD101521\
Data File : VD070717.D
Acq On : 15 Oct 2021 15:17
Operator : VA/SY
Sample : VD1015SBL01
Misc : 5.00G/5.00ml/MSVOA_D/SOIL
ALS Vial : 12 Sample Multiplier: 1

Instrument :
MSVOA_D
ClientSampleId :
VD1015SBL01

Quant Time: Oct 18 08:32:59 2021
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_D\Method\82D101521S.M
Quant Title : SW846 8260
QLast Update : Mon Oct 18 08:32:03 2021
Response via : Initial Calibration

