

Data Path : Z:\voasrv\HPCHEM1\MSVOA_D\Data\VD101723\
 Data File : VD077346.D
 Acq On : 17 Oct 2023 17:50
 Operator : JC/SY
 Sample : VIBLK
 Misc : 5.00G/5.0ml/MSVOA_D/SOIL
 ALS Vial : 11 Sample Multiplier: 1

Instrument :
 MSVOA_D
 ClientSampleId :
 VIBLK

Quant Time: Oct 17 19:37:40 2023
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_D\Method\82D101723S.M
 Quant Title : SW846 8260
 QLast Update : Tue Oct 17 19:32:43 2023
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) Pentafluorobenzene	7.875	168	246579	50.000	ug/l	0.00
34) 1,4-Difluorobenzene	8.781	114	425655	50.000	ug/l	0.00
63) Chlorobenzene-d5	11.581	117	573690	50.000	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	13.522	152	336392	50.000	ug/l	0.00

System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	8.234	65	79596	36.150	ug/l	0.00
Spiked Amount	50.000	Range	50 - 163	Recovery	=	72.300%
35) Dibromofluoromethane	7.810	113	123748	41.382	ug/l	0.00
Spiked Amount	50.000	Range	54 - 147	Recovery	=	82.760%
50) Toluene-d8	10.269	98	535170	49.971	ug/l	0.00
Spiked Amount	50.000	Range	58 - 134	Recovery	=	99.940%
62) 4-Bromofluorobenzene	12.575	95	240126	77.891	ug/l	0.00
Spiked Amount	50.000	Range	30 - 143	Recovery	=	155.780%#

Target Compounds	Qvalue

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_D\Data\VD101723\
Data File : VD077346.D
Acq On : 17 Oct 2023 17:50
Operator : JC/SY
Sample : VIBLK
Misc : 5.00G/5.0ml/MSVOA_D/SOIL
ALS Vial : 11 Sample Multiplier: 1

Instrument :
MSVOA_D
ClientSampleId :
VIBLK

Quant Time: Oct 17 19:37:40 2023
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_D\Method\82D101723S.M
Quant Title : SW846 8260
QLast Update : Tue Oct 17 19:32:43 2023
Response via : Initial Calibration

