

Data Path : Z:\voasrv\HPCHEM1\MSVOA_D\Data\VD101821\
 Data File : VD070725.D
 Acq On : 18 Oct 2021 12:54
 Operator : VA/SY
 Sample : VD1018SBL01
 Misc : 5.00G/5.00ml/MSVOA_D/SOIL
 ALS Vial : 3 Sample Multiplier: 1

Instrument :
 MSVOA_D
 ClientSampleId :
 VD1018SBL01

Quant Time: Oct 19 06:39:44 2021
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_D\Method\82D101521S.M
 Quant Title : SW846 8260
 QLast Update : Mon Oct 18 08:32:03 2021
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) Pentafluorobenzene	7.979	168	488233	50.000	ug/l	0.00
34) 1,4-Difluorobenzene	8.861	114	893304	50.000	ug/l	0.00
63) Chlorobenzene-d5	11.637	117	830538	50.000	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	13.567	152	358518	50.000	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	8.326	65	327149	56.573	ug/l	0.00
Spiked Amount	50.000	Range	50 - 163	Recovery	=	113.140%
35) Dibromofluoromethane	7.914	113	299356	52.336	ug/l	0.00
Spiked Amount	50.000	Range	54 - 147	Recovery	=	104.680%
50) Toluene-d8	10.337	98	1139073	51.950	ug/l	0.00
Spiked Amount	50.000	Range	49 - 140	Recovery	=	103.900%
62) 4-Bromofluorobenzene	12.626	95	367305	48.651	ug/l	0.00
Spiked Amount	50.000	Range	25 - 144	Recovery	=	97.300%
Target Compounds					Qvalue	
95) Naphthalene	15.390	128	4092	5.111	ug/l	96

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_D\Data\VD101821\
Data File : VD070725.D
Acq On : 18 Oct 2021 12:54
Operator : VA/SY
Sample : VD1018SBL01
Misc : 5.00G/5.00ml/MSVOA_D/SOIL
ALS Vial : 3 Sample Multiplier: 1

Instrument :
MSVOA_D
ClientSampleId :
VD1018SBL01

Quant Time: Oct 19 06:39:44 2021
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_D\Method\82D101521S.M
Quant Title : SW846 8260
QLast Update : Mon Oct 18 08:32:03 2021
Response via : Initial Calibration

