

Data Path : Z:\voasrv\HPCHEM1\MSVOA_D\Data\VD101821\
 Data File : VD070734.D
 Acq On : 18 Oct 2021 17:12
 Operator : VA/SY
 Sample : M4259-10
 Misc : 4.96G/5.00ml/MSVOA_D/SOIL
 ALS Vial : 12 Sample Multiplier: 1

Instrument :
 MSVOA_D
 ClientSampleId :
 WC-1-AS

Quant Time: Oct 19 06:41:44 2021
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_D\Method\82D101521S.M
 Quant Title : SW846 8260
 QLast Update : Mon Oct 18 08:32:03 2021
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) Pentafluorobenzene	7.979	168	389502	50.000	ug/l	0.00
34) 1,4-Difluorobenzene	8.861	114	759180	50.000	ug/l	0.00
63) Chlorobenzene-d5	11.638	117	718232	50.000	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	13.561	152	296575	50.000	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	8.326	65	253977	55.052	ug/l	0.00
Spiked Amount	50.000	Range	50 - 163	Recovery	= 110.100%	
35) Dibromofluoromethane	7.914	113	255019	52.461	ug/l	0.00
Spiked Amount	50.000	Range	54 - 147	Recovery	= 104.920%	
50) Toluene-d8	10.338	98	999359	53.630	ug/l	0.00
Spiked Amount	50.000	Range	49 - 140	Recovery	= 107.260%	
62) 4-Bromofluorobenzene	12.626	95	299987	46.754	ug/l	0.00
Spiked Amount	50.000	Range	25 - 144	Recovery	= 93.500%	

Target Compounds	Qvalue

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_D\Data\VD101821\
Data File : VD070734.D
Acq On : 18 Oct 2021 17:12
Operator : VA/SY
Sample : M4259-10
Misc : 4.96G/5.00ml/MSVOA_D/SOIL
ALS Vial : 12 Sample Multiplier: 1

Instrument :
MSVOA_D
ClientSampleId :
WC-1-AS

Quant Time: Oct 19 06:41:44 2021
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_D\Method\82D101521S.M
Quant Title : SW846 8260
QLast Update : Mon Oct 18 08:32:03 2021
Response via : Initial Calibration

