

Data Path : Z:\voasrv\HPCHEM1\MSVOA_D\Data\VD101821\
 Data File : VD070740.D
 Acq On : 18 Oct 2021 20:03
 Operator : VA/SY
 Sample : M4260-01
 Misc : 5.78G/5.00ml/MSVOA_D/SOIL
 ALS Vial : 18 Sample Multiplier: 1

Instrument :
 MSVOA_D
 ClientSampleId :
 TEST-DA-(1)

Quant Time: Oct 19 06:42:47 2021
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_D\Method\82D101521S.M
 Quant Title : SW846 8260
 QLast Update : Mon Oct 18 08:32:03 2021
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) Pentafluorobenzene	7.979	168	363190	50.000	ug/l	0.00
34) 1,4-Difluorobenzene	8.861	114	715783	50.000	ug/l	0.00
63) Chlorobenzene-d5	11.638	117	669932	50.000	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	13.567	152	263124	50.000	ug/l	0.00

System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	8.326	65	234222	54.448	ug/l	0.00
Spiked Amount	50.000	Range	50 - 163	Recovery	=	108.900%
35) Dibromofluoromethane	7.914	113	237784	51.882	ug/l	0.00
Spiked Amount	50.000	Range	54 - 147	Recovery	=	103.760%
50) Toluene-d8	10.338	98	960343	54.661	ug/l	0.00
Spiked Amount	50.000	Range	49 - 140	Recovery	=	109.320%
62) 4-Bromofluorobenzene	12.626	95	273520	45.214	ug/l	0.00
Spiked Amount	50.000	Range	25 - 144	Recovery	=	90.420%

Target Compounds	Qvalue

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_D\Data\VD101821\
Data File : VD070740.D
Acq On : 18 Oct 2021 20:03
Operator : VA/SY
Sample : M4260-01
Misc : 5.78G/5.00ml/MSVOA_D/SOIL
ALS Vial : 18 Sample Multiplier: 1

Instrument :
MSVOA_D
ClientSampleId :
TEST-DA-(1)

Quant Time: Oct 19 06:42:47 2021
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_D\Method\82D101521S.M
Quant Title : SW846 8260
QLast Update : Mon Oct 18 08:32:03 2021
Response via : Initial Calibration

