

Data Path : Z:\voasrv\HPCHEM1\MSVOA_D\Data\VD101821\
 Data File : VD070741.D
 Acq On : 18 Oct 2021 20:32
 Operator : VA/SY
 Sample : M4260-04
 Misc : 5.77G/5.00ml/MSVOA_D/SOIL
 ALS Vial : 19 Sample Multiplier: 1

Instrument :
 MSVOA_D
 ClientSampleId :
 TEST-DB-(1)

Quant Time: Oct 19 06:42:57 2021
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_D\Method\82D101521S.M
 Quant Title : SW846 8260
 QLast Update : Mon Oct 18 08:32:03 2021
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) Pentafluorobenzene	7.979	168	364491	50.000	ug/l	0.00
34) 1,4-Difluorobenzene	8.861	114	728265	50.000	ug/l	0.00
63) Chlorobenzene-d5	11.638	117	697955	50.000	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	13.561	152	285745	50.000	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	8.332	65	253268	58.666	ug/l	0.00
Spiked Amount	50.000	Range	50 - 163	Recovery	=	117.340%
35) Dibromofluoromethane	7.914	113	247501	53.076	ug/l	0.00
Spiked Amount	50.000	Range	54 - 147	Recovery	=	106.160%
50) Toluene-d8	10.332	98	971824	54.367	ug/l	0.00
Spiked Amount	50.000	Range	49 - 140	Recovery	=	108.740%
62) 4-Bromofluorobenzene	12.626	95	293767	47.728	ug/l	0.00
Spiked Amount	50.000	Range	25 - 144	Recovery	=	95.460%

Target Compounds	Qvalue

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_D\Data\VD101821\
Data File : VD070741.D
Acq On : 18 Oct 2021 20:32
Operator : VA/SY
Sample : M4260-04
Misc : 5.77G/5.00ml/MSVOA_D/SOIL
ALS Vial : 19 Sample Multiplier: 1

Instrument :
MSVOA_D
ClientSampleId :
TEST-DB-(1)

Quant Time: Oct 19 06:42:57 2021
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_D\Method\82D101521S.M
Quant Title : SW846 8260
QLast Update : Mon Oct 18 08:32:03 2021
Response via : Initial Calibration

