

Data Path : Z:\voasrv\HPCHEM1\MSVOA_D\Data\VD101821\
 Data File : VD070744.D
 Acq On : 18 Oct 2021 21:57
 Operator : VA/SY
 Sample : M4273-01
 Misc : 5.12G/5.00ml/MSVOA_D/SOIL
 ALS Vial : 22 Sample Multiplier: 1

Instrument :
 MSVOA_D
 ClientSampleId :
 R-LA-4/2-WC

Quant Time: Oct 19 06:43:30 2021
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_D\Method\82D101521S.M
 Quant Title : SW846 8260
 QLast Update : Mon Oct 18 08:32:03 2021
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) Pentafluorobenzene	7.979	168	482000	50.000	ug/l	0.00
34) 1,4-Difluorobenzene	8.861	114	948511	50.000	ug/l	0.00
63) Chlorobenzene-d5	11.638	117	949989	50.000	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	13.561	152	389379	50.000	ug/l	0.00

System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	8.326	65	292883	51.302	ug/l	0.00
Spiked Amount	50.000	Range	50 - 163	Recovery	=	102.600%
35) Dibromofluoromethane	7.908	113	297027	48.906	ug/l	0.00
Spiked Amount	50.000	Range	54 - 147	Recovery	=	97.820%
50) Toluene-d8	10.332	98	1292675	55.524	ug/l	0.00
Spiked Amount	50.000	Range	49 - 140	Recovery	=	111.040%
62) 4-Bromofluorobenzene	12.620	95	444327	55.427	ug/l	0.00
Spiked Amount	50.000	Range	25 - 144	Recovery	=	110.860%

Target Compounds	Qvalue

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_D\Data\VD101821\
Data File : VD070744.D
Acq On : 18 Oct 2021 21:57
Operator : VA/SY
Sample : M4273-01
Misc : 5.12G/5.00ml/MSVOA_D/SOIL
ALS Vial : 22 Sample Multiplier: 1

Instrument :
MSVOA_D
ClientSampleId :
R-LA-4/2-WC

Quant Time: Oct 19 06:43:30 2021
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_D\Method\82D101521S.M
Quant Title : SW846 8260
QLast Update : Mon Oct 18 08:32:03 2021
Response via : Initial Calibration

