

Data Path : Z:\voasrv\HPCHEM1\MSVOA_D\Data\VD101921\
 Data File : VD070748.D
 Acq On : 19 Oct 2021 10:14
 Operator : VA/SY
 Sample : VD1019SBL01
 Misc : 5.00G/5.00ml/MSVOA_D/SOIL
 ALS Vial : 3 Sample Multiplier: 1

Instrument :
 MSVOA_D
 ClientSampleId :
 VD1019SBL01

Quant Time: Oct 20 04:38:51 2021
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_D\Method\82D101521S.M
 Quant Title : SW846 8260
 QLast Update : Mon Oct 18 08:32:03 2021
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) Pentafluorobenzene	7.979	168	444668	50.00	ug/l	0.00
34) 1,4-Difluorobenzene	8.861	114	878954	50.00	ug/l	0.00
63) Chlorobenzene-d5	11.638	117	832915	50.00	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	13.567	152	354394	50.00	ug/l	0.00

System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	8.326	65	328267	62.33	ug/l	0.00
Spiked Amount	50.000	Range	50 - 163	Recovery	=	124.66%
35) Dibromofluoromethane	7.914	113	301582	53.59	ug/l	0.00
Spiked Amount	50.000	Range	54 - 147	Recovery	=	107.18%
50) Toluene-d8	10.338	98	1129752	52.37	ug/l	0.00
Spiked Amount	50.000	Range	49 - 140	Recovery	=	104.74%
62) 4-Bromofluorobenzene	12.626	95	368567	49.61	ug/l	0.00
Spiked Amount	50.000	Range	25 - 144	Recovery	=	99.22%

Target Compounds	Qvalue

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_D\Data\VD101921\
Data File : VD070748.D
Acq On : 19 Oct 2021 10:14
Operator : VA/SY
Sample : VD1019SBL01
Misc : 5.00G/5.00ml/MSVOA_D/SOIL
ALS Vial : 3 Sample Multiplier: 1

Instrument :
MSVOA_D
ClientSampleId :
VD1019SBL01

Quant Time: Oct 20 04:38:51 2021
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_D\Method\82D101521S.M
Quant Title : SW846 8260
QLast Update : Mon Oct 18 08:32:03 2021
Response via : Initial Calibration

