

Data Path : Z:\voasrv\HPCHEM1\MSVOA_D\Data\VD101921\
 Data File : VD070753.D
 Acq On : 19 Oct 2021 12:40
 Operator : VA/SY
 Sample : M4259-04RE
 Misc : 4.83G/5.00ml/MSVOA_D/SOIL
 ALS Vial : 8 Sample Multiplier: 1

Instrument :
 MSVOA_D
 ClientSampleId :
 WC-1-BSRE

Quant Time: Oct 20 04:40:06 2021
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_D\Method\82D101521S.M
 Quant Title : SW846 8260
 QLast Update : Mon Oct 18 08:32:03 2021
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) Pentafluorobenzene	7.973	168	29790	50.00	ug/l	0.00
34) 1,4-Difluorobenzene	8.861	114	47882	50.00	ug/l	0.00
63) Chlorobenzene-d5	11.638	117	52912	50.00	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	13.561	152	21925	50.00	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	8.332	65	33178	94.03	ug/l	0.00
Spiked Amount 50.000	Range 50 - 163		Recovery = 188.06%#			
35) Dibromofluoromethane	7.914	113	20312	66.25	ug/l	0.00
Spiked Amount 50.000	Range 54 - 147		Recovery = 132.50%			
50) Toluene-d8	10.332	98	66042	56.19	ug/l	0.00
Spiked Amount 50.000	Range 49 - 140		Recovery = 112.38%			
62) 4-Bromofluorobenzene	12.620	95	21659	53.52	ug/l	0.00
Spiked Amount 50.000	Range 25 - 144		Recovery = 107.04%			
Target Compounds					Qvalue	
16) Acetone	4.167	43	4425	52.00	ug/l	99

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_D\Data\VD101921\
Data File : VD070753.D
Acq On : 19 Oct 2021 12:40
Operator : VA/SY
Sample : M4259-04RE
Misc : 4.83G/5.00ml/MSVOA_D/SOIL
ALS Vial : 8 Sample Multiplier: 1

Instrument :
MSVOA_D
ClientSampleId :
WC-1-BSRE

Quant Time: Oct 20 04:40:06 2021
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_D\Method\82D101521S.M
Quant Title : SW846 8260
QLast Update : Mon Oct 18 08:32:03 2021
Response via : Initial Calibration

