

Data Path : Z:\voasrv\HPCHEM1\MSVOA_D\Data\VD101921\
 Data File : VD070754.D
 Acq On : 19 Oct 2021 13:08
 Operator : VA/SY
 Sample : M4259-07
 Misc : 3.90G/5.00ml/MSVOA_D/SOIL
 ALS Vial : 9 Sample Multiplier: 1

Instrument :
 MSVOA_D
 ClientSampleId :
 WC-1-BD

Quant Time: Oct 20 04:40:16 2021
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_D\Method\82D101521S.M
 Quant Title : SW846 8260
 QLast Update : Mon Oct 18 08:32:03 2021
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) Pentafluorobenzene	7.979	168	403832	50.00	ug/l	0.00
34) 1,4-Difluorobenzene	8.861	114	783089	50.00	ug/l	0.00
63) Chlorobenzene-d5	11.638	117	767005	50.00	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	13.567	152	331451	50.00	ug/l	0.00

System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	8.332	65	292254	61.10	ug/l	0.00
Spiked Amount	50.000	Range	50 - 163	Recovery	=	122.20%
35) Dibromofluoromethane	7.914	113	274687	54.78	ug/l	0.00
Spiked Amount	50.000	Range	54 - 147	Recovery	=	109.56%
50) Toluene-d8	10.338	98	1049464	54.60	ug/l	0.00
Spiked Amount	50.000	Range	49 - 140	Recovery	=	109.20%
62) 4-Bromofluorobenzene	12.626	95	323267	48.84	ug/l	0.00
Spiked Amount	50.000	Range	25 - 144	Recovery	=	97.68%

Target Compounds	Qvalue

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_D\Data\VD101921\
Data File : VD070754.D
Acq On : 19 Oct 2021 13:08
Operator : VA/SY
Sample : M4259-07
Misc : 3.90G/5.00ml/MSVOA_D/SOIL
ALS Vial : 9 Sample Multiplier: 1

Instrument :
MSVOA_D
ClientSampleId :
WC-1-BD

Quant Time: Oct 20 04:40:16 2021
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_D\Method\82D101521S.M
Quant Title : SW846 8260
QLast Update : Mon Oct 18 08:32:03 2021
Response via : Initial Calibration

