

Data Path : Z:\voasrv\HPCHEM1\MSVOA_D\Data\VD101923\
 Data File : VD077355.D
 Acq On : 19 Oct 2023 15:51
 Operator : JC/SY
 Sample : 04879-03
 Misc : 5.01G/5.0ml/MSVOA_D/SOIL
 ALS Vial : 9 Sample Multiplier: 1

Instrument :
 MSVOA_D
 ClientSampleId :
 TP-2

Quant Time: Oct 20 02:08:17 2023
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_D\Method\82D101723S.M
 Quant Title : SW846 8260
 QLast Update : Wed Oct 18 03:02:43 2023
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) Pentafluorobenzene	7.875	168	140109	50.000	ug/l	0.00
34) 1,4-Difluorobenzene	8.781	114	273629	50.000	ug/l	0.00
63) Chlorobenzene-d5	11.581	117	248722	50.000	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	13.522	152	94584	50.000	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	8.228	65	71975	57.530	ug/l	0.00
Spiked Amount	50.000	Range	50 - 163	Recovery	=	115.060%
35) Dibromofluoromethane	7.804	113	82334	42.830	ug/l	0.00
Spiked Amount	50.000	Range	54 - 147	Recovery	=	85.660%
50) Toluene-d8	10.269	98	264502	38.419	ug/l	0.00
Spiked Amount	50.000	Range	58 - 134	Recovery	=	76.840%
62) 4-Bromofluorobenzene	12.575	95	64404	32.498	ug/l	0.00
Spiked Amount	50.000	Range	30 - 143	Recovery	=	65.000%
Target Compounds						
20) Methylene Chloride	4.799	84	30942	15.326	ug/l	Qvalue 91

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_D\Data\VD101923\
Data File : VD077355.D
Acq On : 19 Oct 2023 15:51
Operator : JC/SY
Sample : 04879-03
Misc : 5.01G/5.0ml/MSVOA_D/SOIL
ALS Vial : 9 Sample Multiplier: 1

Instrument :
MSVOA_D
ClientSampleId :
TP-2

Quant Time: Oct 20 02:08:17 2023
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_D\Method\82D101723S.M
Quant Title : SW846 8260
QLast Update : Wed Oct 18 03:02:43 2023
Response via : Initial Calibration

