

Data Path : Z:\voasrv\HPCHEM1\MSVOA_D\Data\VD102023\
 Data File : VD077371.D
 Acq On : 20 Oct 2023 18:13
 Operator : JC/SY
 Sample : 04980-01
 Misc : 2.73G/5.0ml/MSVOA_D/SOIL
 ALS Vial : 15 Sample Multiplier: 1

Instrument :
 MSVOA_D
 ClientSampleId :
 EXHAUST-VENT

Quant Time: Oct 21 05:31:57 2023
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_D\Method\82D101723S.M
 Quant Title : SW846 8260
 QLast Update : Wed Oct 18 03:02:43 2023
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) Pentafluorobenzene	7.875	168	100074	50.000	ug/l	0.00
34) 1,4-Difluorobenzene	8.775	114	187057	50.000	ug/l	0.00
63) Chlorobenzene-d5	11.581	117	140804	50.000	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	13.516	152	33259	50.000	ug/l	0.00

System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	8.234	65	50713	56.751	ug/l	0.00
Spiked Amount	50.000	Range	50 - 163	Recovery	=	113.500%
35) Dibromofluoromethane	7.804	113	61540	46.828	ug/l	0.00
Spiked Amount	50.000	Range	54 - 147	Recovery	=	93.660%
50) Toluene-d8	10.269	98	161453	34.305	ug/l	0.00
Spiked Amount	50.000	Range	58 - 134	Recovery	=	68.600%
62) 4-Bromofluorobenzene	12.575	95	28862	21.304	ug/l	0.00
Spiked Amount	50.000	Range	30 - 143	Recovery	=	42.600%

Target Compounds	Qvalue

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_D\Data\VD102023\
Data File : VD077371.D
Acq On : 20 Oct 2023 18:13
Operator : JC/SY
Sample : 04980-01
Misc : 2.73G/5.0ml/MSVOA_D/SOIL
ALS Vial : 15 Sample Multiplier: 1

Instrument :
MSVOA_D
ClientSampleId :
EXHAUST-VENT

Quant Time: Oct 21 05:31:57 2023
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_D\Method\82D101723S.M
Quant Title : SW846 8260
QLast Update : Wed Oct 18 03:02:43 2023
Response via : Initial Calibration

