

Data Path : Z:\voasrv\HPCHEM1\MSVOA_D\Data\VD102122\
 Data File : VD074613.D
 Acq On : 21 Oct 2022 14:03
 Operator : VA/SY
 Sample : N5202-01
 Misc : 5.02G/5.00ml/MSVOA_D/SOIL
 ALS Vial : 8 Sample Multiplier: 1

Instrument :
 MSVOA_D
 ClientSampleId :
 SU-04-101922

Quant Time: Oct 21 23:10:36 2022
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_D\Method\82D101922S.M
 Quant Title : SW846 8260
 QLast Update : Thu Oct 20 06:41:20 2022
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) Pentafluorobenzene	7.881	168	115772	50.000	ug/l	0.00
34) 1,4-Difluorobenzene	8.781	114	197360	50.000	ug/l	0.00
63) Chlorobenzene-d5	11.580	117	159160	50.000	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	13.516	152	69673	50.000	ug/l	0.00

System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	8.228	65	49071	55.517	ug/l	0.00
Spiked Amount	50.000	Range	50 - 163	Recovery	=	111.040%
35) Dibromofluoromethane	7.804	113	64665	59.344	ug/l	0.00
Spiked Amount	50.000	Range	54 - 147	Recovery	=	118.680%
50) Toluene-d8	10.269	98	224507	56.834	ug/l	0.00
Spiked Amount	50.000	Range	49 - 140	Recovery	=	113.660%
62) 4-Bromofluorobenzene	12.575	95	60969	53.798	ug/l	0.00
Spiked Amount	50.000	Range	25 - 144	Recovery	=	107.600%

Target Compounds	Qvalue

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_D\Data\VD102122\
Data File : VD074613.D
Acq On : 21 Oct 2022 14:03
Operator : VA/SY
Sample : N5202-01
Misc : 5.02G/5.00ml/MSVOA_D/SOIL
ALS Vial : 8 Sample Multiplier: 1

Instrument :
MSVOA_D
ClientSampleId :
SU-04-101922

Quant Time: Oct 21 23:10:36 2022
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_D\Method\82D101922S.M
Quant Title : SW846 8260
QLast Update : Thu Oct 20 06:41:20 2022
Response via : Initial Calibration

