

Data Path : Z:\voasrv\HPCHEM1\MSVOA_D\Data\VD102122\
 Data File : VD074615.D
 Acq On : 21 Oct 2022 14:58
 Operator : VA/SY
 Sample : N5229-03
 Misc : 5.06G/5.00ml/MSVOA_D/SOIL
 ALS Vial : 10 Sample Multiplier: 1

Instrument :
 MSVOA_D
 ClientSampleId :
 LRH-COMP-2

Quant Time: Oct 21 23:10:58 2022
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_D\Method\82D101922S.M
 Quant Title : SW846 8260
 QLast Update : Thu Oct 20 06:41:20 2022
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) Pentafluorobenzene	7.875	168	60207	50.000	ug/l	0.00
34) 1,4-Difluorobenzene	8.775	114	105665	50.000	ug/l	0.00
63) Chlorobenzene-d5	11.581	117	65675	50.000	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	13.516	152	11243	50.000	ug/l	0.00

System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	8.228	65	34226	74.458	ug/l	0.00
Spiked Amount	50.000	Range	50 - 163	Recovery	=	148.920%
35) Dibromofluoromethane	7.805	113	42018	72.264	ug/l	0.00
Spiked Amount	50.000	Range	54 - 147	Recovery	=	144.520%
50) Toluene-d8	10.269	98	115807	54.757	ug/l	0.00
Spiked Amount	50.000	Range	49 - 140	Recovery	=	109.520%
62) 4-Bromofluorobenzene	12.575	95	15398	24.313	ug/l	0.00
Spiked Amount	50.000	Range	25 - 144	Recovery	=	48.620%

Target Compounds	Qvalue

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_D\Data\VD102122\
Data File : VD074615.D
Acq On : 21 Oct 2022 14:58
Operator : VA/SY
Sample : N5229-03
Misc : 5.06G/5.00ml/MSVOA_D/SOIL
ALS Vial : 10 Sample Multiplier: 1

Instrument :
MSVOA_D
ClientSampleId :
LRH-COMP-2

Quant Time: Oct 21 23:10:58 2022
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_D\Method\82D101922S.M
Quant Title : SW846 8260
QLast Update : Thu Oct 20 06:41:20 2022
Response via : Initial Calibration

