

Data Path : Z:\voasrv\HPCHEM1\MSVOA_D\Data\VD102122\
 Data File : VD074619.D
 Acq On : 21 Oct 2022 16:49
 Operator : VA/SY
 Sample : N5218-21
 Misc : 5.03G/5.00ml/MSVOA_D/SOIL
 ALS Vial : 14 Sample Multiplier: 1

Instrument :
 MSVOA_D
 ClientSampleId :
 SP-3

Quant Time: Oct 21 23:11:34 2022
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_D\Method\82D101922S.M
 Quant Title : SW846 8260
 QLast Update : Thu Oct 20 06:41:20 2022
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) Pentafluorobenzene	7.875	168	92990	50.000	ug/l	0.00
34) 1,4-Difluorobenzene	8.781	114	155590	50.000	ug/l	0.00
63) Chlorobenzene-d5	11.581	117	120796	50.000	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	13.516	152	46808	50.000	ug/l	0.00

System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	8.228	65	42966	60.519	ug/l	0.00
Spiked Amount	50.000	Range	50 - 163	Recovery	=	121.040%
35) Dibromofluoromethane	7.804	113	57540	67.127	ug/l	0.00
Spiked Amount	50.000	Range	54 - 147	Recovery	=	134.260%
50) Toluene-d8	10.269	98	186678	59.945	ug/l	0.00
Spiked Amount	50.000	Range	49 - 140	Recovery	=	119.880%
62) 4-Bromofluorobenzene	12.569	95	46948	52.500	ug/l	0.00
Spiked Amount	50.000	Range	25 - 144	Recovery	=	105.000%

Target Compounds	Qvalue

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_D\Data\VD102122\
Data File : VD074619.D
Acq On : 21 Oct 2022 16:49
Operator : VA/SY
Sample : N5218-21
Misc : 5.03G/5.00ml/MSVOA_D/SOIL
ALS Vial : 14 Sample Multiplier: 1

Instrument :
MSVOA_D
ClientSampleId :
SP-3

Quant Time: Oct 21 23:11:34 2022
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_D\Method\82D101922S.M
Quant Title : SW846 8260
QLast Update : Thu Oct 20 06:41:20 2022
Response via : Initial Calibration

