

Data Path : Z:\voasrv\HPCHEM1\MSVOA_D\Data\VD102221\
 Data File : VD070818.D
 Acq On : 22 Oct 2021 11:57
 Operator : VA/SY
 Sample : M4297-01
 Misc : 5.09G/5.00ml/MSVOA_D/SOIL
 ALS Vial : 6 Sample Multiplier: 1

Instrument :
 MSVOA_D
 ClientSampleId :
 TR-04-10202021

Quant Time: Oct 23 04:10:09 2021
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_D\Method\82D101521S.M
 Quant Title : SW846 8260
 QLast Update : Mon Oct 18 08:32:03 2021
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) Pentafluorobenzene	7.973	168	440996	50.000	ug/l	0.00
34) 1,4-Difluorobenzene	8.861	114	796344	50.000	ug/l	0.00
63) Chlorobenzene-d5	11.638	117	729177	50.000	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	13.561	152	266649	50.000	ug/l	0.00

System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	8.326	65	289040	55.337	ug/l	0.00
Spiked Amount	50.000	Range	50 - 163	Recovery	=	110.680%
35) Dibromofluoromethane	7.908	113	262279	51.437	ug/l	0.00
Spiked Amount	50.000	Range	54 - 147	Recovery	=	102.880%
50) Toluene-d8	10.332	98	1036443	53.025	ug/l	0.00
Spiked Amount	50.000	Range	49 - 140	Recovery	=	106.040%
62) 4-Bromofluorobenzene	12.620	95	311472	46.278	ug/l	0.00
Spiked Amount	50.000	Range	25 - 144	Recovery	=	92.560%

Target Compounds	Qvalue

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_D\Data\VD102221\
Data File : VD070818.D
Acq On : 22 Oct 2021 11:57
Operator : VA/SY
Sample : M4297-01
Misc : 5.09G/5.00ml/MSVOA_D/SOIL
ALS Vial : 6 Sample Multiplier: 1

Instrument :
MSVOA_D
ClientSampleId :
TR-04-10202021

Quant Time: Oct 23 04:10:09 2021
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_D\Method\82D101521S.M
Quant Title : SW846 8260
QLast Update : Mon Oct 18 08:32:03 2021
Response via : Initial Calibration

