

Data Path : Z:\voasrv\HPCHEM1\MSVOA_D\Data\VD102221\
 Data File : VD070825.D
 Acq On : 22 Oct 2021 15:16
 Operator : VA/SY
 Sample : M4290-07
 Misc : 5.02G/5.00ml/MSVOA_D/SOIL
 ALS Vial : 13 Sample Multiplier: 1

Instrument :
 MSVOA_D
 ClientSampleId :
 TH-BD

Quant Time: Oct 23 04:11:51 2021
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_D\Method\82D101521S.M
 Quant Title : SW846 8260
 QLast Update : Mon Oct 18 08:32:03 2021
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) Pentafluorobenzene	7.973	168	46455	50.000	ug/l	0.00
34) 1,4-Difluorobenzene	8.861	114	80900	50.000	ug/l	0.00
63) Chlorobenzene-d5	11.638	117	89414	50.000	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	13.561	152	37633	50.000	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	8.326	65	52771	95.908	ug/l	0.00
Spiked Amount	50.000	Range	50 - 163	Recovery	=	191.820%#
35) Dibromofluoromethane	7.914	113	33581	64.827	ug/l	0.00
Spiked Amount	50.000	Range	54 - 147	Recovery	=	129.660%
50) Toluene-d8	10.332	98	105461	53.110	ug/l	0.00
Spiked Amount	50.000	Range	49 - 140	Recovery	=	106.220%
62) 4-Bromofluorobenzene	12.620	95	38139	55.780	ug/l	0.00
Spiked Amount	50.000	Range	25 - 144	Recovery	=	111.560%
Target Compounds						
16) Acetone	4.156	43	4801	36.176	ug/l	Qvalue # 78

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_D\Data\VD102221\
Data File : VD070825.D
Acq On : 22 Oct 2021 15:16
Operator : VA/SY
Sample : M4290-07
Misc : 5.02G/5.00ml/MSVOA_D/SOIL
ALS Vial : 13 Sample Multiplier: 1

Instrument :
MSVOA_D
ClientSampleId :
TH-BD

Quant Time: Oct 23 04:11:51 2021
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_D\Method\82D101521S.M
Quant Title : SW846 8260
QLast Update : Mon Oct 18 08:32:03 2021
Response via : Initial Calibration

