

Data Path : Z:\voasrv\HPCHEM1\MSVOA_D\Data\VD102221\
 Data File : VD070826.D
 Acq On : 22 Oct 2021 15:45
 Operator : VA/SY
 Sample : M4290-10
 Misc : 4.29G/5.00ml/MSVOA_D/SOIL
 ALS Vial : 14 Sample Multiplier: 1

Instrument :
 MSVOA_D
 ClientSampleId :
 TH-BBS

Quant Time: Oct 23 04:12:06 2021
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_D\Method\82D101521S.M
 Quant Title : SW846 8260
 QLast Update : Mon Oct 18 08:32:03 2021
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) Pentafluorobenzene	7.979	168	99887	50.000	ug/l	0.00
34) 1,4-Difluorobenzene	8.861	114	176830	50.000	ug/l	0.00
63) Chlorobenzene-d5	11.638	117	183906	50.000	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	13.561	152	83721	50.000	ug/l	0.00

System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	8.332	65	109684	92.710	ug/l	0.00
Spiked Amount	50.000	Range	50 - 163	Recovery	=	185.420%#
35) Dibromofluoromethane	7.914	113	77442	68.396	ug/l	0.00
Spiked Amount	50.000	Range	54 - 147	Recovery	=	136.800%
50) Toluene-d8	10.332	98	230242	53.047	ug/l	0.00
Spiked Amount	50.000	Range	49 - 140	Recovery	=	106.100%
62) 4-Bromofluorobenzene	12.620	95	81584	54.590	ug/l	0.00
Spiked Amount	50.000	Range	25 - 144	Recovery	=	109.180%

Target Compounds	Qvalue

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_D\Data\VD102221\
Data File : VD070826.D
Acq On : 22 Oct 2021 15:45
Operator : VA/SY
Sample : M4290-10
Misc : 4.29G/5.00ml/MSVOA_D/SOIL
ALS Vial : 14 Sample Multiplier: 1

Instrument :
MSVOA_D
ClientSampleId :
TH-BBS

Quant Time: Oct 23 04:12:06 2021
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_D\Method\82D101521S.M
Quant Title : SW846 8260
QLast Update : Mon Oct 18 08:32:03 2021
Response via : Initial Calibration

