

Data Path : Z:\voasrv\HPCHEM1\MSVOA_D\Data\VD102423\
 Data File : VD077394.D
 Acq On : 24 Oct 2023 15:00
 Operator : JC/SY
 Sample : 05011-01
 Misc : 5.08G/5.0ml/MSVOA_D/SOIL
 ALS Vial : 11 Sample Multiplier: 1

Instrument :
 MSVOA_D
 ClientSampleId :
 RBR-200027

Quant Time: Oct 25 04:05:23 2023
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_D\Method\82D101723S.M
 Quant Title : SW846 8260
 QLast Update : Wed Oct 18 03:02:43 2023
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) Pentafluorobenzene	7.875	168	85654	50.000	ug/l	0.00
34) 1,4-Difluorobenzene	8.781	114	160205	50.000	ug/l	0.00
63) Chlorobenzene-d5	11.581	117	68151	50.000	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	13.522	152	24952	50.000	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	8.228	65	46761	61.138	ug/l	0.00
Spiked Amount	50.000	Range	50 - 163	Recovery	=	122.280%
35) Dibromofluoromethane	7.810	113	53374	47.422	ug/l	0.00
Spiked Amount	50.000	Range	54 - 147	Recovery	=	94.840%
50) Toluene-d8	10.269	98	37475	9.297	ug/l	0.00
Spiked Amount	50.000	Range	58 - 134	Recovery	=	18.600%#
62) 4-Bromofluorobenzene	12.569	95	18908	16.296	ug/l	0.00
Spiked Amount	50.000	Range	30 - 143	Recovery	=	32.600%
Target Compounds						Qvalue
11) Tert butyl alcohol	5.051	59	4308	100.128	ug/l #	77
18) Methyl Acetate	4.569	43	4776	8.182	ug/l #	88

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_D\Data\VD102423\
Data File : VD077394.D
Acq On : 24 Oct 2023 15:00
Operator : JC/SY
Sample : 05011-01
Misc : 5.08G/5.0ml/MSVOA_D/SOIL
ALS Vial : 11 Sample Multiplier: 1

Instrument :
MSVOA_D
ClientSampleId :
RBR-200027

Quant Time: Oct 25 04:05:23 2023
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_D\Method\82D101723S.M
Quant Title : SW846 8260
QLast Update : Wed Oct 18 03:02:43 2023
Response via : Initial Calibration

