



Data Path : Z:\voasrv\HPCHEM1\MSVOA_D\Data\VD102423\
Data File : VD077387.D
Acq On : 24 Oct 2023 09:29
Operator : JC/SY
Sample : BFB
Misc : 5.00G/5.0ml/MSVOA_D/SOIL
ALS Vial : 1 Sample Multiplier: 1

Instrument :
MSVOA_D
ClientSampleId :
BFB

Integration File: RTEINT.P

Method : Z:\voasrv\HPCHEM1\MSVOA_D\Method\82D101723S.M
Title : SW846 8260
Last Update : Wed Oct 18 03:02:43 2023



AutoFind: Scans 1867, 1868, 1869; Background Corrected with Scan 1859

Target Mass	Rel. to Mass	Lower Limit%	Upper Limit%	Rel. Abn%	Raw Abn	Result Pass/Fail
50	95	15	40	17.2	13367	PASS
75	95	30	60	50.6	39293	PASS
95	95	100	100	100.0	77725	PASS
96	95	5	9	6.4	4979	PASS
173	174	0.00	2	1.6	1183	PASS
174	95	50	100	94.6	73547	PASS
175	174	5	9	7.3	5371	PASS
176	174	95	101	96.0	70616	PASS
177	176	5	9	6.9	4850	PASS