

Data Path : Z:\voasrv\HPCHEM1\MSVOA_D\Data\VD102523\
 Data File : VD077407.D
 Acq On : 25 Oct 2023 11:41
 Operator : JC/SY
 Sample : VD1025SBL01
 Misc : 5.00G/5.0ml/MSVOA_D/SOIL
 ALS Vial : 5 Sample Multiplier: 1

Instrument :
 MSVOA_D
 ClientSampleId :
 VD1025SBL01

Quant Time: Oct 26 06:57:06 2023
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_D\Method\82D101723S.M
 Quant Title : SW846 8260
 QLast Update : Wed Oct 18 03:02:43 2023
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) Pentafluorobenzene	7.881	168	179340	50.000	ug/l	0.00
34) 1,4-Difluorobenzene	8.775	114	334270	50.000	ug/l	0.00
63) Chlorobenzene-d5	11.581	117	329960	50.000	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	13.522	152	126008	50.000	ug/l	0.00

System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	8.234	65	82375	51.439	ug/l	0.00
Spiked Amount	50.000	Range	50 - 163	Recovery	=	102.880%
35) Dibromofluoromethane	7.804	113	87953	37.452	ug/l	0.00
Spiked Amount	50.000	Range	54 - 147	Recovery	=	74.900%
50) Toluene-d8	10.269	98	292973	34.835	ug/l	0.00
Spiked Amount	50.000	Range	58 - 134	Recovery	=	69.660%
62) 4-Bromofluorobenzene	12.575	95	79769	32.949	ug/l	0.00
Spiked Amount	50.000	Range	30 - 143	Recovery	=	65.900%

Target Compounds	Qvalue

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_D\Data\VD102523\
Data File : VD077407.D
Acq On : 25 Oct 2023 11:41
Operator : JC/SY
Sample : VD1025SBL01
Misc : 5.00G/5.0ml/MSVOA_D/SOIL
ALS Vial : 5 Sample Multiplier: 1

Instrument :
MSVOA_D
ClientSampleId :
VD1025SBL01

Quant Time: Oct 26 06:57:06 2023
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_D\Method\82D101723S.M
Quant Title : SW846 8260
QLast Update : Wed Oct 18 03:02:43 2023
Response via : Initial Calibration

