

Data Path : Z:\voasrv\HPCHEM1\MSVOA_D\Data\VD102523\
 Data File : VD077410.D
 Acq On : 25 Oct 2023 13:25
 Operator : JC/SY
 Sample : 05025-01RE
 Misc : 5.08G/5.0ml/MSVOA_D/SOIL
 ALS Vial : 8 Sample Multiplier: 1

Instrument :
 MSVOA_D
 ClientSampleId :
 EO-01-10232023-ARE

Quant Time: Oct 25 16:11:50 2023
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_D\Method\82D101723S.M
 Quant Title : SW846 8260
 QLast Update : Wed Oct 18 03:02:43 2023
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) Pentafluorobenzene	7.875	168	81445	50.000	ug/l	0.00
34) 1,4-Difluorobenzene	8.775	114	158415	50.000	ug/l	0.00
63) Chlorobenzene-d5	11.581	117	133529	50.000	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	13.516	152	41803	50.000	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	8.234	65	55643	76.511	ug/l	0.00
Spiked Amount	50.000	Range	50 - 163	Recovery	=	153.020%
35) Dibromofluoromethane	7.804	113	56494	50.761	ug/l	0.00
Spiked Amount	50.000	Range	54 - 147	Recovery	=	101.520%
50) Toluene-d8	10.269	98	126311	31.690	ug/l	0.00
Spiked Amount	50.000	Range	58 - 134	Recovery	=	63.380%
62) 4-Bromofluorobenzene	12.575	95	33047	28.803	ug/l	0.00
Spiked Amount	50.000	Range	30 - 143	Recovery	=	57.600%
Target Compounds						
20) Methylene Chloride	4.799	84	11425	7.811	ug/l	91

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_D\Data\VD102523\
Data File : VD077410.D
Acq On : 25 Oct 2023 13:25
Operator : JC/SY
Sample : 05025-01RE
Misc : 5.08G/5.0ml/MSVOA_D/SOIL
ALS Vial : 8 Sample Multiplier: 1

Instrument :
MSVOA_D
ClientSampleId :
EO-01-10232023-ARE

Quant Time: Oct 25 16:11:50 2023
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_D\Method\82D101723S.M
Quant Title : SW846 8260
QLast Update : Wed Oct 18 03:02:43 2023
Response via : Initial Calibration

