

Data Path : Z:\voasrv\HPCHEM1\MSVOA_D\Data\VD102523\
 Data File : VD077413.D
 Acq On : 25 Oct 2023 14:46
 Operator : JC/SY
 Sample : 05011-05RE
 Misc : 5.04G/5.0ml/MSVOA_D/SOIL
 ALS Vial : 11 Sample Multiplier: 1

Instrument :
 MSVOA_D
 ClientSampleId :
 RBR-24313RE

Quant Time: Oct 26 06:59:56 2023
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_D\Method\82D101723S.M
 Quant Title : SW846 8260
 QLast Update : Wed Oct 18 03:02:43 2023
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) Pentafluorobenzene	7.881	168	145207	50.000	ug/l	0.00
34) 1,4-Difluorobenzene	8.775	114	261796	50.000	ug/l	0.00
63) Chlorobenzene-d5	11.581	117	162496	50.000	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	13.516	152	43838	50.000	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	8.228	65	76061	58.661	ug/l	0.00
Spiked Amount	50.000	Range	50 - 163	Recovery	=	117.320%
35) Dibromofluoromethane	7.804	113	77631	42.208	ug/l	0.00
Spiked Amount	50.000	Range	54 - 147	Recovery	=	84.420%
50) Toluene-d8	10.269	98	173812	26.388	ug/l	0.00
Spiked Amount	50.000	Range	58 - 134	Recovery	=	52.780%#
62) 4-Bromofluorobenzene	12.575	95	40618	21.422	ug/l	0.00
Spiked Amount	50.000	Range	30 - 143	Recovery	=	42.840%
Target Compounds						
					Qvalue	
20) Methylene Chloride	4.810	84	21776	8.715	ug/l	96

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_D\Data\VD102523\
Data File : VD077413.D
Acq On : 25 Oct 2023 14:46
Operator : JC/SY
Sample : 05011-05RE
Misc : 5.04G/5.0ml/MSVOA_D/SOIL
ALS Vial : 11 Sample Multiplier: 1

Instrument :
MSVOA_D
ClientSampleId :
RBR-24313RE

Quant Time: Oct 26 06:59:56 2023
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_D\Method\82D101723S.M
Quant Title : SW846 8260
QLast Update : Wed Oct 18 03:02:43 2023
Response via : Initial Calibration

