

Data Path : Z:\voasrv\HPCHEM1\MSVOA_D\Data\VD102523\
 Data File : VD077417.D
 Acq On : 25 Oct 2023 16:34
 Operator : JC/SY
 Sample : 05021-01
 Misc : 5.05G/5.0ml/MSVOA_D/SOIL
 ALS Vial : 15 Sample Multiplier: 1

Instrument :
 MSVOA_D
 ClientSampleId :
 350

Quant Time: Oct 26 07:01:26 2023
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_D\Method\82D101723S.M
 Quant Title : SW846 8260
 QLast Update : Wed Oct 18 03:02:43 2023
 Response via : Initial Calibration

Compound	R.T. QIon		Response	Conc	Units	Dev(Min)

Internal Standards						
1) Pentafluorobenzene	7.881	168	15746	50.000	ug/l	0.00
34) 1,4-Difluorobenzene	8.775	114	34756	50.000	ug/l	0.00
63) Chlorobenzene-d5	11.587	117	39450	50.000	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	13.522	152	17593	50.000	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	8.228	65	14902	105.987	ug/l	0.00
Spiked Amount	50.000	Range 50 - 163	Recovery	=	211.980	%#
35) Dibromofluoromethane	7.805	113	13565	55.554	ug/l	0.00
Spiked Amount	50.000	Range 54 - 147	Recovery	=	111.100	%
50) Toluene-d8	10.269	98	29092	33.268	ug/l	0.00
Spiked Amount	50.000	Range 58 - 134	Recovery	=	66.540	%
62) 4-Bromofluorobenzene	12.581	95	10889	43.258	ug/l	0.00
Spiked Amount	50.000	Range 30 - 143	Recovery	=	86.520	%
Target Compounds						
20) Methylene Chloride	4.799	84	10678	57.983	ug/l	Qvalue 95

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_D\Data\VD102523\
Data File : VD077417.D
Acq On : 25 Oct 2023 16:34
Operator : JC/SY
Sample : 05021-01
Misc : 5.05G/5.0ml/MSVOA_D/SOIL
ALS Vial : 15 Sample Multiplier: 1

Instrument :
MSVOA_D
ClientSampleId :
350

Quant Time: Oct 26 07:01:26 2023
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_D\Method\82D101723S.M
Quant Title : SW846 8260
QLast Update : Wed Oct 18 03:02:43 2023
Response via : Initial Calibration

