

Data Path : Z:\voasrv\HPCHEM1\MSVOA_D\Data\VD102621\
 Data File : VD070847.D
 Acq On : 26 Oct 2021 18:28
 Operator : VA/SY
 Sample : M4348-05
 Misc : 5.06G/5.00ml/MSVOA_D/SOIL
 ALS Vial : 20 Sample Multiplier: 1

Instrument :
 MSVOA_D
 ClientSampleId :
 VNJ-229

Quant Time: Oct 27 06:23:23 2021
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_D\Method\82D102621S.M
 Quant Title : SW846 8260
 QLast Update : Wed Oct 27 05:34:47 2021
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) Pentafluorobenzene	7.979	168	401831	50.000	ug/l	0.00
34) 1,4-Difluorobenzene	8.861	114	738824	50.000	ug/l	0.00
63) Chlorobenzene-d5	11.638	117	712754	50.000	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	13.561	152	323312	50.000	ug/l	0.00

System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	8.326	65	224902	55.649	ug/l	0.00
Spiked Amount	50.000	Range	50 - 163	Recovery	=	111.300%
35) Dibromofluoromethane	7.909	113	249129	51.141	ug/l	0.00
Spiked Amount	50.000	Range	54 - 147	Recovery	=	102.280%
50) Toluene-d8	10.332	98	951458	51.915	ug/l	0.00
Spiked Amount	50.000	Range	49 - 140	Recovery	=	103.840%
62) 4-Bromofluorobenzene	12.620	95	297094	49.131	ug/l	0.00
Spiked Amount	50.000	Range	25 - 144	Recovery	=	98.260%

Target Compounds	Qvalue

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_D\Data\VD102621\
Data File : VD070847.D
Acq On : 26 Oct 2021 18:28
Operator : VA/SY
Sample : M4348-05
Misc : 5.06G/5.00ml/MSVOA_D/SOIL
ALS Vial : 20 Sample Multiplier: 1

Instrument :
MSVOA_D
ClientSampleId :
VNJ-229

Quant Time: Oct 27 06:23:23 2021
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_D\Method\82D102621S.M
Quant Title : SW846 8260
QLast Update : Wed Oct 27 05:34:47 2021
Response via : Initial Calibration

