

Data Path : Z:\voasrv\HPCHEM1\MSVOA_D\Data\VD102623\
 Data File : VD077427.D
 Acq On : 26 Oct 2023 12:02
 Operator : JC/SY
 Sample : VD1026SBL01
 Misc : 5.00G/5.0ml/MSVOA_D/SOIL
 ALS Vial : 5 Sample Multiplier: 1

Instrument :
 MSVOA_D
 ClientSampleId :
 VD1026

Quant Time: Oct 27 07:46:18 2023
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_D\Method\82D101723S.M
 Quant Title : SW846 8260
 QLast Update : Wed Oct 18 03:02:43 2023
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) Pentafluorobenzene	7.875	168	144915	50.000	ug/l	0.00
34) 1,4-Difluorobenzene	8.775	114	305061	50.000	ug/l	0.00
63) Chlorobenzene-d5	11.581	117	321886	50.000	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	13.516	152	152160	50.000	ug/l	0.00

System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	8.228	65	65956	50.970	ug/l	0.00
Spiked Amount	50.000	Range	50 - 163	Recovery	=	101.940%
35) Dibromofluoromethane	7.805	113	75535	35.244	ug/l	0.00
Spiked Amount	50.000	Range	54 - 147	Recovery	=	70.480%
50) Toluene-d8	10.269	98	350909	45.718	ug/l	0.00
Spiked Amount	50.000	Range	58 - 134	Recovery	=	91.440%
62) 4-Bromofluorobenzene	12.575	95	97592	44.171	ug/l	0.00
Spiked Amount	50.000	Range	30 - 143	Recovery	=	88.340%

Target Compounds	Qvalue

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_D\Data\VD102623\
Data File : VD077427.D
Acq On : 26 Oct 2023 12:02
Operator : JC/SY
Sample : VD1026SBL01
Misc : 5.00G/5.0ml/MSVOA_D/SOIL
ALS Vial : 5 Sample Multiplier: 1

Instrument :
MSVOA_D
ClientSampleId :
VD1026

Quant Time: Oct 27 07:46:18 2023
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_D\Method\82D101723S.M
Quant Title : SW846 8260
QLast Update : Wed Oct 18 03:02:43 2023
Response via : Initial Calibration

