

Data Path : Z:\voasrv\HPCHEM1\MSVOA_D\Data\VD102821\
 Data File : VD070896.D
 Acq On : 28 Oct 2021 17:19
 Operator : VA/SY
 Sample : M4395-04
 Misc : 2.85G/5.00ml/MSVOA_D/SOIL
 ALS Vial : 15 Sample Multiplier: 1

Instrument :
 MSVOA_D
 ClientSampleId :
 MH-FFD

Quant Time: Oct 29 06:35:51 2021
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_D\Method\82D102621S.M
 Quant Title : SW846 8260
 QLast Update : Wed Oct 27 05:34:47 2021
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) Pentafluorobenzene	7.979	168	651064	50.000	ug/l	0.00
34) 1,4-Difluorobenzene	8.861	114	1105892	50.000	ug/l	0.00
63) Chlorobenzene-d5	11.638	117	1022679	50.000	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	13.561	152	459162	50.000	ug/l	0.00

System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	8.326	65	268592	41.018	ug/l	0.00
Spiked Amount	50.000	Range	50 - 163	Recovery	=	82.040%
35) Dibromofluoromethane	7.908	113	333924	45.795	ug/l	0.00
Spiked Amount	50.000	Range	54 - 147	Recovery	=	91.600%
50) Toluene-d8	10.332	98	1367306	49.842	ug/l	0.00
Spiked Amount	50.000	Range	49 - 140	Recovery	=	99.680%
62) 4-Bromofluorobenzene	12.620	95	438389	48.434	ug/l	0.00
Spiked Amount	50.000	Range	25 - 144	Recovery	=	96.860%

Target Compounds	Qvalue

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_D\Data\VD102821\
Data File : VD070896.D
Acq On : 28 Oct 2021 17:19
Operator : VA/SY
Sample : M4395-04
Misc : 2.85G/5.00ml/MSVOA_D/SOIL
ALS Vial : 15 Sample Multiplier: 1

Instrument :
MSVOA_D
ClientSampleId :
MH-FFD

Quant Time: Oct 29 06:35:51 2021
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_D\Method\82D102621S.M
Quant Title : SW846 8260
QLast Update : Wed Oct 27 05:34:47 2021
Response via : Initial Calibration

