

Data Path : Z:\VOASRV\HPCHEM1\MSVOA D\DATA\VD102920\
 Data File : VD067458.D
 Acq On : 29 Oct 2020 18:03
 Operator : VA/SY
 Sample : L4552-01
 Misc : 5.00G/5.00ml/MSVOA D/SOIL
 ALS Vial : 15 Sample Multiplier: 1

Instrument :
 MSVOA_D
 ClientSampleId :
 VW-STONE-A

Quant Time: Oct 30 02:29:55 2020
 Quant Method : Z:\VOASRV\HPCHEM1\MSVOA_D\METHOD\82D101320S.M
 Quant Title : SW846 8260
 QLast Update : Mon Oct 19 19:13:29 2020
 Response via : Initial Calibration

Internal Standards	R.T.	QIon	Response	Conc	Units	Dev(Min)
1) Pentafluorobenzene	0.00	168	0	0.00	ug/l	-7.98
34) 1,4-Difluorobenzene	8.87	114	214	50.00	ug/l	0.00
63) Chlorobenzene-d5	11.64	117	233	50.00	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	13.58	152	170	50.00	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	0.00	65	0	0.00	ug/l	
Spiked Amount	50.000		Recovery	=	0.00%	
35) Dibromofluoromethane	0.00	113	0	0.00	ug/l	
Spiked Amount	50.000		Recovery	=	0.00%	
50) Toluene-d8	0.00	98	0	0.00	ug/l	
Spiked Amount	50.000		Recovery	=	0.00%	
62) 4-Bromofluorobenzene	0.00	95	0	0.00	ug/l	
Spiked Amount	50.000		Recovery	=	0.00%	
Target Compounds						
43) Isopropyl Acetate	8.38	43	67	63.273	ug/l #	45
59) 2-Hexanone	10.78	43	110	304.794	ug/l #	22

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\V0ASRV\HPCHEM1\MSVOA D\DATA\VD102920\
Data File : VD067458.D
Acq On : 29 Oct 2020 18:03
Operator : VA/SY
Sample : L4552-01
Misc : 5.00G/5.00ml/MSVOA D/SOIL
ALS Vial : 15 Sample Multiplier: 1

Instrument :
MSVOA_D
ClientSampleId :
VW-STONE-A

Quant Time: Oct 30 02:29:55 2020
Quant Method : Z:\V0ASRV\HPCHEM1\MSVOA_D\METHOD\82D101320S.M
Quant Title : SW846 8260
QLast Update : Mon Oct 19 19:13:29 2020
Response via : Initial Calibration

