

Data Path : Z:\VOASRV\HPCHEM1\MSVOA D\DATA\VD103118\
 Data File : VD060245.D
 Acq On : 31 Oct 2018 16:52
 Operator : VA/AP
 Sample : J5199-07
 Misc : 6.35g/5ml/MSVOA D/SOIL
 ALS Vial : 7 Sample Multiplier: 1

Instrument :
 MSVOA_D
 ClientSampleId :
 OK-01-103018

Quant Time: Nov 01 00:30:21 2018
 Quant Method : Z:\VOASRV\HPCHEM1\MSVOA_D\METHOD\82D100118S.M
 Quant Title : SW846 8260
 QLast Update : Tue Oct 02 06:56:35 2018
 Response via : Initial Calibration

Internal Standards	R.T.	QIon	Response	Conc	Units	Dev(Min)
1) Pentafluorobenzene	6.42	168	1687155	50.00	ug/l	0.03
34) 1,4-Difluorobenzene	7.45	114	2449854	50.00	ug/l	0.03
63) Chlorobenzene-d5	11.29	117	1665199	50.00	ug/l	0.02
72) 1,4-Dichlorobenzene-d4	13.37	152	579807	50.00	ug/l	0.01
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	6.79	65	631361	49.37	ug/l	0.03
Spiked Amount 50.000			Recovery	=	98.74%	
35) Dibromofluoromethane	6.33	113	787652	42.34	ug/l	0.03
Spiked Amount 50.000			Recovery	=	84.68%	
50) Toluene-d8	9.46	98	1896485	35.75	ug/l	0.02
Spiked Amount 50.000			Recovery	=	71.50%	
62) 4-Bromofluorobenzene	12.43	95	571988	28.80	ug/l	0.01
Spiked Amount 50.000			Recovery	=	57.60%	
Target Compounds						
8) Diethyl Ether	2.83	74	7191	2.494	ug/l	71
20) Methylene Chloride	3.72	84	135488	3.410	ug/l	97

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\V0ASRV\HPCHEM1\MSVOA D\DATA\VD103118\
Data File : VD060245.D
Acq On : 31 Oct 2018 16:52
Operator : VA/AP
Sample : J5199-07
Misc : 6.35g/5ml/MSVOA D/SOIL
ALS Vial : 7 Sample Multiplier: 1

Instrument :
MSVOA_D
ClientSampleId :
OK-01-103018

Quant Time: Nov 01 00:30:21 2018
Quant Method : Z:\V0ASRV\HPCHEM1\MSVOA_D\METHOD\82D100118S.M
Quant Title : SW846 8260
QLast Update : Tue Oct 02 06:56:35 2018
Response via : Initial Calibration

