

Data Path : Z:\voasrv\HPCHEM1\MSVOA_D\Data\VD103122\
 Data File : VD074689.D
 Acq On : 31 Oct 2022 15:41
 Operator : VA/SY
 Sample : N5303-01
 Misc : 5.01G/5.00ml/MSVOA_D/SOIL
 ALS Vial : 8 Sample Multiplier: 1

Instrument :
 MSVOA_D
 ClientSampleId :
 NWB-1839

Quant Time: Nov 01 01:33:44 2022
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_D\Method\82D102722S.M
 Quant Title : SW846 8260
 QLast Update : Fri Oct 28 00:18:17 2022
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) Pentafluorobenzene	7.875	168	137290	50.000	ug/l	0.00
34) 1,4-Difluorobenzene	8.781	114	249272	50.000	ug/l	0.00
63) Chlorobenzene-d5	11.581	117	208103	50.000	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	13.516	152	94573	50.000	ug/l	0.00

System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	8.228	65	76636	66.137	ug/l	0.00
Spiked Amount	50.000	Range	50 - 163	Recovery	=	132.280%
35) Dibromofluoromethane	7.804	113	90442	59.419	ug/l	0.00
Spiked Amount	50.000	Range	54 - 147	Recovery	=	118.840%
50) Toluene-d8	10.269	98	289760	49.501	ug/l	0.00
Spiked Amount	50.000	Range	49 - 140	Recovery	=	99.000%
62) 4-Bromofluorobenzene	12.575	95	84395	52.497	ug/l	0.00
Spiked Amount	50.000	Range	25 - 144	Recovery	=	105.000%

Target Compounds	Qvalue

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_D\Data\VD103122\
Data File : VD074689.D
Acq On : 31 Oct 2022 15:41
Operator : VA/SY
Sample : N5303-01
Misc : 5.01G/5.00ml/MSVOA_D/SOIL
ALS Vial : 8 Sample Multiplier: 1

Instrument :
MSVOA_D
ClientSampleId :
NWB-1839

Quant Time: Nov 01 01:33:44 2022
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_D\Method\82D102722S.M
Quant Title : SW846 8260
QLast Update : Fri Oct 28 00:18:17 2022
Response via : Initial Calibration

