

Data Path : Z:\voasrv\HPCHEM1\MSVOA_D\Data\VD103122\
 Data File : VD074695.D
 Acq On : 31 Oct 2022 18:27
 Operator : VA/SY
 Sample : N5341-01
 Misc : 5.01G/5.00ml/MSVOA_D/SOIL
 ALS Vial : 14 Sample Multiplier: 1

Instrument :
 MSVOA_D
 ClientSampleId :
 WC-1

Quant Time: Nov 01 01:35:09 2022
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_D\Method\82D102722S.M
 Quant Title : SW846 8260
 QLast Update : Fri Oct 28 00:18:17 2022
 Response via : Initial Calibration

Compound	R.T. QIon		Response	Conc	Units	Dev(Min)

Internal Standards						
1) Pentafluorobenzene	7.881	168	16194	50.000	ug/l	0.00
34) 1,4-Difluorobenzene	8.781	114	32321	50.000	ug/l	0.00
63) Chlorobenzene-d5	11.581	117	31788	50.000	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	13.516	152	15969	50.000	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	8.228	65	18183	133.033	ug/l	0.00
Spiked Amount	50.000	Range	50 - 163	Recovery	=	266.060%#
35) Dibromofluoromethane	7.805	113	16161	81.886	ug/l	0.00
Spiked Amount	50.000	Range	54 - 147	Recovery	=	163.780%#
50) Toluene-d8	10.269	98	41000	54.020	ug/l	0.00
Spiked Amount	50.000	Range	49 - 140	Recovery	=	108.040%
62) 4-Bromofluorobenzene	12.569	95	15284	73.324	ug/l	0.00
Spiked Amount	50.000	Range	25 - 144	Recovery	=	146.640%#
Target Compounds						
						Qvalue
20) Methylene Chloride	4.799	84	6354	22.054	ug/l #	84

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_D\Data\VD103122\
Data File : VD074695.D
Acq On : 31 Oct 2022 18:27
Operator : VA/SY
Sample : N5341-01
Misc : 5.01G/5.00ml/MSVOA_D/SOIL
ALS Vial : 14 Sample Multiplier: 1

Instrument :
MSVOA_D
ClientSampleId :
WC-1

Quant Time: Nov 01 01:35:09 2022
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_D\Method\82D102722S.M
Quant Title : SW846 8260
QLast Update : Fri Oct 28 00:18:17 2022
Response via : Initial Calibration

