

Data Path : Z:\voasrv\HPCHEM1\MSVOA_D\Data\VD103122\
 Data File : VD074697.D
 Acq On : 31 Oct 2022 19:23
 Operator : VA/SY
 Sample : VSTDCCC050
 Misc : 5.00G/5.00ml/MSVOA_D/SOIL
 ALS Vial : 16 Sample Multiplier: 1

Instrument :
 MSVOA_D
 LabSampleId :
 VSTDCCC050

Quant Time: Nov 01 01:35:26 2022
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_D\Method\82D102722S.M
 Quant Title : SW846 8260
 QLast Update : Fri Oct 28 00:18:17 2022
 Response via : Initial Calibration

Min. RRF : 0.000 Min. Rel. Area : 50% Max. R.T. Dev 0.50min
 Max. RRF Dev : 25% Max. Rel. Area : 150%

	Compound	AvgRF	CCRF	%Dev	Area%	Dev(min)
1 I	Pentafluorobenzene	1.000	1.000	0.0	92	0.00
2 T	Dichlorodifluoromethane	0.431	0.373	13.5	86	0.00
3 P	Chloromethane	0.670	0.544	18.8	81	0.00
4 C	Vinyl Chloride	0.714	0.657	8.0#	87	0.00
5 T	Bromomethane	0.583	0.458	21.4	85	0.00
6 T	Chloroethane	0.486	0.464	4.5	92	0.00
7 T	Trichlorofluoromethane	0.830	0.817	1.6	92	0.00
8 T	Diethyl Ether	0.251	0.242	3.6	91	0.00
9 T	1,1,2-Trichlorotrifluoroeth	0.514	0.537	-4.5	97	0.00
10 T	Methyl Iodide	0.592	0.590	0.3	90	0.00
11 T	Tert butyl alcohol	0.058	0.038	34.5#	103	0.00
12 CM	1,1-Dichloroethene	0.545	0.520	4.6#	91	0.00
13 T	Acrolein	0.037	0.032	13.5	83	0.00
14 T	Allyl chloride	0.566	0.564	0.4	93	0.00
15 T	Acrylonitrile	0.097	0.101	-4.1	102	0.00
16 T	Acetone	0.078	0.073	6.4	83	0.00
17 T	Carbon Disulfide	1.653	1.298	21.5	74	0.00
18 T	Methyl Acetate	0.243	0.225	7.4	101	0.00
19 T	Methyl tert-butyl Ether	1.045	1.102	-5.5	98	0.00
20 T	Methylene Chloride	1.057	0.694	34.3#	96	0.00
21 T	trans-1,2-Dichloroethene	0.605	0.583	3.6	89	0.00
22 T	Diisopropyl ether	1.191	1.305	-9.6	99	0.00
23 T	Vinyl Acetate	0.463	0.488	-5.4	95	0.00
24 P	1,1-Dichloroethane	0.896	0.924	-3.1	96	0.00
25 T	2-Butanone	0.107	0.108	-0.9	94	0.00
26 T	2,2-Dichloropropane	0.784	0.827	-5.5	98	0.00
27 T	cis-1,2-Dichloroethene	0.658	0.676	-2.7	96	0.00
28 T	Bromochloromethane	0.308	0.322	-4.5	98	0.00
29 T	Tetrahydrofuran	0.065	0.067	-3.1	100	0.00
30 C	Chloroform	0.964	1.041	-8.0#	102	0.00
31 T	Cyclohexane	0.773	0.736	4.8	87	0.00
32 T	1,1,1-Trichloroethane	0.847	0.910	-7.4	99	0.00
33 S	1,2-Dichloroethane-d4	0.422	0.413	2.1	92	0.00
34 I	1,4-Difluorobenzene	1.000	1.000	0.0	92	0.00
35 S	Dibromofluoromethane	0.305	0.319	-4.6	95	0.00
36 T	1,1-Dichloropropene	0.435	0.458	-5.3	92	0.00
37 T	Ethyl Acetate	0.139	0.157	-12.9	106	0.00
38 T	Carbon Tetrachloride	0.404	0.439	-8.7	91	0.00
39 T	Methylcyclohexane	0.532	0.511	3.9	81	0.00
40 TM	Benzene	1.327	1.392	-4.9	93	0.00
41 T	Methacrylonitrile	0.072	0.083	-15.3	99	0.00
42 TM	1,2-Dichloroethane	0.312	0.329	-5.4	96	0.00
43 T	Isopropyl Acetate	0.268	0.290	-8.2	101	0.00
44 TM	Trichloroethene	0.389	0.353	9.3	81	0.00
45 C	1,2-Dichloropropane	0.296	0.294	0.7#	89	0.00
46 T	Dibromomethane	0.173	0.169	2.3	89	0.00
47 T	Bromodichloromethane	0.416	0.414	0.5	89	0.00
48 T	Methyl methacrylate	0.119	0.122	-2.5	93	0.00

Data Path : Z:\voasrv\HPCHEM1\MSVOA_D\Data\VD103122\
 Data File : VD074697.D
 Acq On : 31 Oct 2022 19:23
 Operator : VA/SY
 Sample : VSTDCCC050
 Misc : 5.00G/5.00ml/MSVOA_D/SOIL
 ALS Vial : 16 Sample Multiplier: 1

Instrument :
 MSVOA_D
 LabSampleId :
 VSTDCCC050

Quant Time: Nov 01 01:35:26 2022
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_D\Method\82D102722S.M
 Quant Title : SW846 8260
 QLast Update : Fri Oct 28 00:18:17 2022
 Response via : Initial Calibration

Min. RRF : 0.000 Min. Rel. Area : 50% Max. R.T. Dev 0.50min
 Max. RRF Dev : 25% Max. Rel. Area : 150%

	Compound	AvgRF	CCRF	%Dev	Area%	Dev(min)
49 T	1,4-Dioxane	0.002	0.002	0.0	81	0.00
50 S	Toluene-d8	1.174	1.114	5.1	81	0.00
51 T	4-Methyl-2-Pentanone	0.129	0.139	-7.8	94	0.00
52 CM	Toluene	0.831	0.828	0.4#	84	0.00
53 T	t-1,3-Dichloropropene	0.357	0.373	-4.5	90	0.00
54 T	cis-1,3-Dichloropropene	0.455	0.466	-2.4	88	0.00
55 T	1,1,2-Trichloroethane	0.236	0.240	-1.7	91	0.00
56 T	Ethyl methacrylate	0.259	0.272	-5.0	90	0.00
57 T	1,3-Dichloropropane	0.371	0.385	-3.8	91	0.00
58 T	2-Chloroethyl Vinyl ether	0.054	0.057	-5.6	91	0.00
59 T	2-Hexanone	0.088	0.093	-5.7	90	0.00
60 T	Dibromochloromethane	0.280	0.284	-1.4	86	0.00
61 T	1,2-Dibromoethane	0.220	0.219	0.5	87	0.00
62 S	4-Bromofluorobenzene	0.322	0.328	-1.9	93	0.00
63 I	Chlorobenzene-d5	1.000	1.000	0.0	83	0.00
64 T	Tetrachloroethene	0.354	0.360	-1.7	84	0.00
65 PM	Chlorobenzene	1.013	1.080	-6.6	86	0.00
66 T	1,1,1,2-Tetrachloroethane	0.342	0.374	-9.4	88	0.00
67 C	Ethyl Benzene	1.751	1.946	-11.1#	88	0.00
68 T	m/p-Xylenes	0.691	0.769	-11.3	90	0.00
69 T	o-Xylene	0.637	0.713	-11.9	95	0.00
70 T	Styrene	1.072	1.231	-14.8	97	0.00
71 P	Bromoform	0.173	0.198	-14.5	100	0.00
72 I	1,4-Dichlorobenzene-d4	1.000	1.000	0.0	95	0.00
73 T	Isopropylbenzene	3.565	3.974	-11.5	98	0.00
74 T	N-amyl acetate	0.562	0.637	-13.3	103	0.00
75 P	1,1,2,2-Tetrachloroethane	0.593	0.651	-9.8	103	0.00
76 T	1,2,3-Trichloropropane	0.391	0.413	-5.6	89	0.00
77 T	Bromobenzene	0.852	0.885	-3.9	95	0.00
78 T	n-propylbenzene	4.278	4.823	-12.7	99	0.00
79 T	2-Chlorotoluene	2.336	2.545	-8.9	99	0.00
80 T	1,3,5-Trimethylbenzene	2.962	3.245	-9.6	97	0.00
81 T	trans-1,4-Dichloro-2-butene	0.148	0.161	-8.8	99	0.00
82 T	4-Chlorotoluene	2.411	2.669	-10.7	99	0.00
83 T	tert-Butylbenzene	2.628	2.927	-11.4	98	0.00
84 T	1,2,4-Trimethylbenzene	2.922	3.286	-12.5	99	0.00
85 T	sec-Butylbenzene	3.857	4.342	-12.6	99	0.00
86 T	p-Isopropyltoluene	3.297	3.686	-11.8	99	0.00
87 T	1,3-Dichlorobenzene	1.763	1.856	-5.3	97	0.00
88 T	1,4-Dichlorobenzene	1.737	1.807	-4.0	98	0.00
89 T	n-Butylbenzene	2.987	3.375	-13.0	99	0.00
90 T	Hexachloroethane	0.562	0.632	-12.5	100	0.00
91 T	1,2-Dichlorobenzene	1.495	1.590	-6.4	98	0.00
92 T	1,2-Dibromo-3-Chloropropane	0.079	0.088	-11.4	101	0.00
93 T	1,2,4-Trichlorobenzene	0.959	0.993	-3.5	93	0.00
94 T	Hexachlorobutadiene	0.490	0.521	-6.3	94	0.00
95 T	Naphthalene	1.762	1.881	-6.8	96	0.00

Data Path : Z:\voasrv\HPCHEM1\MSVOA_D\Data\VD103122\
Data File : VD074697.D
Acq On : 31 Oct 2022 19:23
Operator : VA/SY
Sample : VSTDCCC050
Misc : 5.00G/5.00ml/MSVOA_D/SOIL
ALS Vial : 16 Sample Multiplier: 1

Instrument :
MSVOA_D
LabSampleId :
VSTDCCC050

Quant Time: Nov 01 01:35:26 2022
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_D\Method\82D102722S.M
Quant Title : SW846 8260
QLast Update : Fri Oct 28 00:18:17 2022
Response via : Initial Calibration

Min. RRF : 0.000 Min. Rel. Area : 50% Max. R.T. Dev 0.50min
Max. RRF Dev : 25% Max. Rel. Area : 150%

	Compound	AvgRF	CCRF	%Dev	Area%	Dev(min)
96 T	1,2,3-Trichlorobenzene	0.826	0.861	-4.2	95	0.00

(#) = Out of Range

SPCC's out = 0 CCC's out = 6