

Data Path : Z:\voasrv\HPCHEM1\MSVOA_D\Data\VD103123\
 Data File : VD077453.D
 Acq On : 31 Oct 2023 11:16
 Operator : JC/SY
 Sample : VD1031SBL01
 Misc : 5.00G/5.0ml/MSVOA_D/SOIL
 ALS Vial : 5 Sample Multiplier: 1

Instrument :
 MSVOA_D
 ClientSampleId :
 VD1031SBL01

Quant Time: Nov 01 06:02:41 2023
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_D\Method\82D103023S.M
 Quant Title : SW846 8260
 QLast Update : Tue Oct 31 04:19:17 2023
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) Pentafluorobenzene	7.881	168	298590	50.000	ug/l	# 0.00
34) 1,4-Difluorobenzene	8.781	114	533784	50.000	ug/l	0.00
63) Chlorobenzene-d5	11.581	117	466634	50.000	ug/l	# 0.00
72) 1,4-Dichlorobenzene-d4	13.522	152	204400	50.000	ug/l	0.00

System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	8.234	65	202387	47.934	ug/l	0.00
Spiked Amount	50.000	Range	50 - 163	Recovery	=	95.860%
35) Dibromofluoromethane	7.804	113	178226	48.309	ug/l	0.00
Spiked Amount	50.000	Range	54 - 147	Recovery	=	96.620%
50) Toluene-d8	10.269	98	714344	49.849	ug/l	0.00
Spiked Amount	50.000	Range	58 - 134	Recovery	=	99.700%
62) 4-Bromofluorobenzene	12.575	95	227530	47.142	ug/l	0.00
Spiked Amount	50.000	Range	30 - 143	Recovery	=	94.280%

Target Compounds	Qvalue

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_D\Data\VD103123\
Data File : VD077453.D
Acq On : 31 Oct 2023 11:16
Operator : JC/SY
Sample : VD1031SBL01
Misc : 5.00G/5.0ml/MSVOA_D/SOIL
ALS Vial : 5 Sample Multiplier: 1

Instrument :
MSVOA_D
ClientSampleId :
VD1031SBL01

Quant Time: Nov 01 06:02:41 2023
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_D\Method\82D103023S.M
Quant Title : SW846 8260
QLast Update : Tue Oct 31 04:19:17 2023
Response via : Initial Calibration

