

Data Path : Z:\voasrv\HPCHEM1\MSVOA_D\Data\VD103123\
 Data File : VD077464.D
 Acq On : 31 Oct 2023 16:42
 Operator : JC/SY
 Sample : 05162-01
 Misc : 5.01G/5.0ml/MSVOA_D/SOIL
 ALS Vial : 16 Sample Multiplier: 1

Instrument :
 MSVOA_D
 ClientSampleId :
 HD-01-10302023

Quant Time: Nov 01 06:09:12 2023
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_D\Method\82D103023S.M
 Quant Title : SW846 8260
 QLast Update : Tue Oct 31 04:19:17 2023
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) Pentafluorobenzene	7.881	168	177177	50.000	ug/l	# 0.00
34) 1,4-Difluorobenzene	8.781	114	399236	50.000	ug/l	0.00
63) Chlorobenzene-d5	11.581	117	458237	50.000	ug/l	# 0.00
72) 1,4-Dichlorobenzene-d4	13.516	152	241994	50.000	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	8.234	65	146570	58.502	ug/l	0.00
Spiked Amount	50.000	Range	50 - 163	Recovery	=	117.000%
35) Dibromofluoromethane	7.805	113	132446	47.999	ug/l	0.00
Spiked Amount	50.000	Range	54 - 147	Recovery	=	96.000%
50) Toluene-d8	10.269	98	535457	49.958	ug/l	0.00
Spiked Amount	50.000	Range	58 - 134	Recovery	=	99.920%
62) 4-Bromofluorobenzene	12.575	95	242254	67.108	ug/l	0.00
Spiked Amount	50.000	Range	30 - 143	Recovery	=	134.220%
Target Compounds						
16) Acetone	4.028	43	21943	33.099	ug/l	# 77

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_D\Data\VD103123\
Data File : VD077464.D
Acq On : 31 Oct 2023 16:42
Operator : JC/SY
Sample : 05162-01
Misc : 5.01G/5.0ml/MSVOA_D/SOIL
ALS Vial : 16 Sample Multiplier: 1

Instrument :
MSVOA_D
ClientSampleId :
HD-01-10302023

Quant Time: Nov 01 06:09:12 2023
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_D\Method\82D103023S.M
Quant Title : SW846 8260
QLast Update : Tue Oct 31 04:19:17 2023
Response via : Initial Calibration

