

Data Path : Z:\voasrv\HPCHEM1\MSVOA_D\Data\VD103123\
 Data File : VD077468.D
 Acq On : 31 Oct 2023 18:32
 Operator : JC/SY
 Sample : 05164-01
 Misc : 3.45G/5.0ml/MSVOA_D/SOIL
 ALS Vial : 20 Sample Multiplier: 1

Instrument :
 MSVOA_D
 ClientSampleId :
 ETGI-352

Quant Time: Nov 01 06:10:46 2023
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_D\Method\82D103023S.M
 Quant Title : SW846 8260
 QLast Update : Tue Oct 31 04:19:17 2023
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) Pentafluorobenzene	7.875	168	265013	50.000	ug/l	# 0.00
34) 1,4-Difluorobenzene	8.781	114	573537	50.000	ug/l	0.00
63) Chlorobenzene-d5	11.581	117	597847	50.000	ug/l	# 0.00
72) 1,4-Dichlorobenzene-d4	13.516	152	288737	50.000	ug/l	0.00

System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	8.234	65	180914	48.277	ug/l	0.00
Spiked Amount	50.000	Range	50 - 163	Recovery	=	96.560%
35) Dibromofluoromethane	7.804	113	182088	45.935	ug/l	0.00
Spiked Amount	50.000	Range	54 - 147	Recovery	=	91.860%
50) Toluene-d8	10.269	98	772807	50.191	ug/l	0.00
Spiked Amount	50.000	Range	58 - 134	Recovery	=	100.380%
62) 4-Bromofluorobenzene	12.575	95	294242	56.738	ug/l	0.00
Spiked Amount	50.000	Range	30 - 143	Recovery	=	113.480%

Target Compounds	Qvalue

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_D\Data\VD103123\
Data File : VD077468.D
Acq On : 31 Oct 2023 18:32
Operator : JC/SY
Sample : 05164-01
Misc : 3.45G/5.0ml/MSVOA_D/SOIL
ALS Vial : 20 Sample Multiplier: 1

Instrument :
MSVOA_D
ClientSampleId :
ETGI-352

Quant Time: Nov 01 06:10:46 2023
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_D\Method\82D103023S.M
Quant Title : SW846 8260
QLast Update : Tue Oct 31 04:19:17 2023
Response via : Initial Calibration

