

Data Path : Z:\voasrv\HPCHEM1\MSVOA_D\Data\VD103123\
 Data File : VD077453.D
 Acq On : 31 Oct 2023 11:16
 Operator : JC/SY
 Sample : VD1031SBL01
 Misc : 5.00G/5.0ml/MSVOA_D/SOIL
 ALS Vial : 5 Sample Multiplier: 1

Instrument :
 MSVOA_D
 ClientSampleId :
 VD1031SBL01

Quant Time: Nov 01 06:02:41 2023
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_D\Method\82D103023S.M
 Quant Title : SW846 8260
 QLast Update : Tue Oct 31 04:19:17 2023
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) Pentafluorobenzene	7.881	168	298590	50.000	ug/l	# 0.00
34) 1,4-Difluorobenzene	8.781	114	533784	50.000	ug/l	0.00
63) Chlorobenzene-d5	11.581	117	466634	50.000	ug/l	# 0.00
72) 1,4-Dichlorobenzene-d4	13.522	152	204400	50.000	ug/l	0.00

System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	8.234	65	202387	47.934	ug/l	0.00
Spiked Amount	50.000	Range	50 - 163	Recovery	=	95.860%
35) Dibromofluoromethane	7.804	113	178226	48.309	ug/l	0.00
Spiked Amount	50.000	Range	54 - 147	Recovery	=	96.620%
50) Toluene-d8	10.269	98	714344	49.849	ug/l	0.00
Spiked Amount	50.000	Range	58 - 134	Recovery	=	99.700%
62) 4-Bromofluorobenzene	12.575	95	227530	47.142	ug/l	0.00
Spiked Amount	50.000	Range	30 - 143	Recovery	=	94.280%


Target Compounds	Qvalue


(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_D\Data\VD103123\
Data File : VD077453.D
Acq On : 31 Oct 2023 11:16
Operator : JC/SY
Sample : VD1031SBL01
Misc : 5.00G/5.0ml/MSVOA_D/SOIL
ALS Vial : 5 Sample Multiplier: 1

Instrument :
MSVOA_D
ClientSampleId :
VD1031SBL01

Quant Time: Nov 01 06:02:41 2023
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_D\Method\82D103023S.M
Quant Title : SW846 8260
QLast Update : Tue Oct 31 04:19:17 2023
Response via : Initial Calibration









#1
Pentafluorobenzene
Concen: 50.000 ug/l
RT: 7.881 min Scan# 110
Delta R.T. 0.006 min
Lab File: VD077453.D
Acq: 31 Oct 2023 11:16




Instrument :
MSVOA_D
ClientSampleId :
VD1031SBL01


Tgt Ion:168 Resp: 298590
Ion Ratio Lower Upper
168 100
99 60.0 36.7 55.1#



#33
1,2-Dichloroethane-d4
Concen: 47.934 ug/l
RT: 8.234 min Scan# 1130
Delta R.T. 0.006 min
Lab File: VD077453.D
Acq: 31 Oct 2023 11:16

Tgt Ion: 65 Resp: 202387
Ion Ratio Lower Upper
65 100
67 53.4 0.0 115.0





#34

1,4-Difluorobenzene

Concen: 50.000 ug/l

RT: 8.781 min Scan# 1

Delta R.T. 0.006 min

Lab File: VD077453.D

Acq: 31 Oct 2023 11:16

Instrument :

MSVOA_D

ClientSampleId :

VD1031SBL01





Tgt Ion:114 Resp: 533784

Ion Ratio Lower Upper

114 100

63 24.1 0.0 33.4

88 16.8 0.0 26.2



#35

Dibromofluoromethane

Concen: 48.309 ug/l

RT: 7.804 min Scan# 1057

Delta R.T. -0.000 min

Lab File: VD077453.D

Acq: 31 Oct 2023 11:16




Tgt Ion:113 Resp: 178226


Ion Ratio Lower Upper

113 100

111 102.8 84.7 127.1

192 18.6 16.8 25.2









#50
Toluene-d8
Concen: 49.849 ug/l
RT: 10.269 min Scan# 1476
Delta R.T. -0.000 min
Lab File: VD077453.D
Acq: 31 Oct 2023 11:16




Instrument :
MSVOA_D
ClientSampleId :
VD1031SBL01


Tgt Ion: 98 Resp: 714344
Ion Ratio Lower Upper
98 100
100 62.9 50.3 75.5



#62
4-Bromofluorobenzene
Concen: 47.142 ug/l
RT: 12.575 min Scan# 1868
Delta R.T. -0.000 min
Lab File: VD077453.D
Acq: 31 Oct 2023 11:16

Tgt Ion: 95 Resp: 227530
Ion Ratio Lower Upper
95 100
174 74.9 0.0 201.8
176 74.6 0.0 191.8









#63
Chlorobenzene-d5
Concen: 50.000 ug/l
RT: 11.581 min Scan# 11
Delta R.T. -0.000 min
Lab File: VD077453.D
Acq: 31 Oct 2023 11:16

Instrument :
MSVOA_D
ClientSampleId :
VD1031SBL01

Tgt Ion:117 Resp: 466634
Ion Ratio Lower Upper
117 100
82 60.8 38.3 57.5#
119 31.8 25.0 37.4



#72
1,4-Dichlorobenzene-d4
Concen: 50.000 ug/l
RT: 13.522 min Scan# 2029
Delta R.T. 0.006 min
Lab File: VD077453.D
Acq: 31 Oct 2023 11:16

Tgt Ion:152 Resp: 204400
Ion Ratio Lower Upper
152 100
115 62.1 27.1 81.3
150 156.9 0.0 353.6

