

Data Path : Z:\voasrv\HPCHEM1\MSVOA_D\Data\VD103123\
 Data File : VD077461.D
 Acq On : 31 Oct 2023 15:20
 Operator : JC/SY
 Sample : 05141-04
 Misc : 5.01G/5.0ml/MSVOA_D/SOIL
 ALS Vial : 13 Sample Multiplier: 1

Instrument :
 MSVOA_D
 ClientSampleId :
 PE-4

Quant Time: Nov 01 06:08:09 2023
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_D\Method\82D103023S.M
 Quant Title : SW846 8260
 QLast Update : Tue Oct 31 04:19:17 2023
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) Pentafluorobenzene	7.881	168	263454	50.000	ug/l	# 0.00
34) 1,4-Difluorobenzene	8.781	114	541216	50.000	ug/l	0.00
63) Chlorobenzene-d5	11.581	117	598205	50.000	ug/l	# 0.00
72) 1,4-Dichlorobenzene-d4	13.516	152	299592	50.000	ug/l	0.00

System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	8.234	65	153939	41.322	ug/l	0.00
Spiked Amount	50.000	Range	50 - 163	Recovery	=	82.640%
35) Dibromofluoromethane	7.811	113	161880	43.276	ug/l	0.00
Spiked Amount	50.000	Range	54 - 147	Recovery	=	86.560%
50) Toluene-d8	10.269	98	710957	48.931	ug/l	0.00
Spiked Amount	50.000	Range	58 - 134	Recovery	=	97.860%
62) 4-Bromofluorobenzene	12.575	95	309241	63.192	ug/l	0.00
Spiked Amount	50.000	Range	30 - 143	Recovery	=	126.380%


Target Compounds	Qvalue


(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_D\Data\VD103123\
Data File : VD077461.D
Acq On : 31 Oct 2023 15:20
Operator : JC/SY
Sample : 05141-04
Misc : 5.01G/5.0ml/MSVOA_D/SOIL
ALS Vial : 13 Sample Multiplier: 1

Instrument :
MSVOA_D
ClientSampleId :
PE-4

Quant Time: Nov 01 06:08:09 2023
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_D\Method\82D103023S.M
Quant Title : SW846 8260
QLast Update : Tue Oct 31 04:19:17 2023
Response via : Initial Calibration









#1
Pentafluorobenzene
Concen: 50.000 ug/l
RT: 7.881 min Scan# 1107
Delta R.T. 0.006 min
Lab File: VD077461.D
Acq: 31 Oct 2023 15:20




Instrument :
MSVOA_D
ClientSampleId :
PE-4


Tgt Ion:168 Resp: 263454
Ion Ratio Lower Upper
168 100
99 60.0 36.7 55.1#



#33
1,2-Dichloroethane-d4
Concen: 41.322 ug/l
RT: 8.234 min Scan# 1130
Delta R.T. 0.006 min
Lab File: VD077461.D
Acq: 31 Oct 2023 15:20

Tgt Ion: 65 Resp: 153939
Ion Ratio Lower Upper
65 100
67 53.8 0.0 115.0





#34

1,4-Difluorobenzene

Concen: 50.000 ug/l

RT: 8.781 min Scan# 1

Delta R.T. 0.006 min

Lab File: VD077461.D

Acq: 31 Oct 2023 15:20

Instrument :

MSVOA_D

ClientSampleId :

PE-4





Tgt Ion:114 Resp: 541216

Ion Ratio Lower Upper

114 100

63 24.2 0.0 33.4

88 16.0 0.0 26.2



#35

Dibromofluoromethane

Concen: 43.276 ug/l

RT: 7.811 min Scan# 1058

Delta R.T. 0.006 min

Lab File: VD077461.D

Acq: 31 Oct 2023 15:20




Tgt Ion:113 Resp: 161880


Ion Ratio Lower Upper

113 100

111 103.9 84.7 127.1

192 18.5 16.8 25.2









#50
Toluene-d8
Concen: 48.931 ug/l
RT: 10.269 min Scan# 1476
Delta R.T. 0.000 min
Lab File: VD077461.D
Acq: 31 Oct 2023 15:20




Instrument :
MSVOA_D
ClientSampleId :
PE-4


Tgt Ion: 98 Resp: 710957
Ion Ratio Lower Upper
98 100
100 63.5 50.3 75.5



#62
4-Bromofluorobenzene
Concen: 63.192 ug/l
RT: 12.575 min Scan# 1868
Delta R.T. 0.000 min
Lab File: VD077461.D
Acq: 31 Oct 2023 15:20

Tgt Ion: 95 Resp: 309241
Ion Ratio Lower Upper
95 100
174 75.4 0.0 201.8
176 72.6 0.0 191.8





#63

Chlorobenzene-d5

Concen: 50.000 ug/l

RT: 11.581 min Scan# 1

Delta R.T. 0.000 min

Lab File: VD077461.D

Acq: 31 Oct 2023 15:20

Instrument :

MSVOA_D

ClientSampleId :

PE-4





Tgt Ion:117 Resp: 598205

Ion Ratio Lower Upper

117 100

82 61.0 38.3 57.5#

119 31.9 25.0 37.4



#72

1,4-Dichlorobenzene-d4

Concen: 50.000 ug/l

RT: 13.516 min Scan# 2028

Delta R.T. 0.000 min

Lab File: VD077461.D

Acq: 31 Oct 2023 15:20

Tgt Ion:152 Resp: 299592

Ion Ratio Lower Upper

152 100

115 62.1 27.1 81.3

150 158.1 0.0 353.6

