



Data Path : Z:\voasrv\HPCHEM1\MSVOA_D\Data\VD110123\
Data File : VD077471.D
Acq On : 01 Nov 2023 08:55
Operator : JC/SY
Sample : BFB
Misc : 5.00G/5.0ml/MSVOA_D/SOIL
ALS Vial : 1 Sample Multiplier: 1

Instrument :
MSVOA_D
ClientSampleId :
BFB

Integration File: RTEINT.P

Method : Z:\voasrv\HPCHEM1\MSVOA_D\Method\82D103023S.M
Title : SW846 8260
Last Update : Tue Oct 31 04:19:17 2023



AutoFind: Scans 1867, 1868, 1869; Background Corrected with Scan 1859

Target Mass	Rel. to Mass	Lower Limit%	Upper Limit%	Rel. Abn%	Raw Abn	Result Pass/Fail
50	95	15	40	24.6	44019	PASS
75	95	30	60	57.6	103200	PASS
95	95	100	100	100.0	179157	PASS
96	95	5	9	6.9	12429	PASS
173	174	0.00	2	1.5	2095	PASS
174	95	50	100	78.7	141013	PASS
175	174	5	9	8.1	11387	PASS
176	174	95	101	96.9	136677	PASS
177	176	5	9	7.1	9699	PASS