

Data Path : Z:\voasrv\HPCHEM1\MSVOA_D\Data\VD110223\
 Data File : VD077491.D
 Acq On : 02 Nov 2023 15:44
 Operator : JC/SY
 Sample : 05213-01
 Misc : 5.01G/5.0ml/MSVOA_D/SOIL
 ALS Vial : 8 Sample Multiplier: 1

Instrument :
 MSVOA_D
 ClientSampleId :
 ETGI-293

Quant Time: Nov 03 01:04:05 2023
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_D\Method\82D103023S.M
 Quant Title : SW846 8260
 QLast Update : Tue Oct 31 04:19:17 2023
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) Pentafluorobenzene	7.881	168	11090	50.000	ug/l	# 0.00
34) 1,4-Difluorobenzene	8.781	114	23019	50.000	ug/l	0.00
63) Chlorobenzene-d5	11.581	117	24265	50.000	ug/l	# 0.00
72) 1,4-Dichlorobenzene-d4	13.516	152	10985	50.000	ug/l	0.00

System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	8.234	65	12066	76.943	ug/l	0.00
Spiked Amount	50.000	Range	50 - 163	Recovery	=	153.880%
35) Dibromofluoromethane	7.805	113	8588	53.979	ug/l	0.00
Spiked Amount	50.000	Range	54 - 147	Recovery	=	107.960%
50) Toluene-d8	10.275	98	31523	51.010	ug/l	0.00
Spiked Amount	50.000	Range	58 - 134	Recovery	=	102.020%
62) 4-Bromofluorobenzene	12.575	95	11633	55.890	ug/l	0.00
Spiked Amount	50.000	Range	30 - 143	Recovery	=	111.780%

Target Compounds	Qvalue

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_D\Data\VD110223\
Data File : VD077491.D
Acq On : 02 Nov 2023 15:44
Operator : JC/SY
Sample : 05213-01
Misc : 5.01G/5.0ml/MSVOA_D/SOIL
ALS Vial : 8 Sample Multiplier: 1

Instrument :
MSVOA_D
ClientSampleId :
ETGI-293

Quant Time: Nov 03 01:04:05 2023
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_D\Method\82D103023S.M
Quant Title : SW846 8260
QLast Update : Tue Oct 31 04:19:17 2023
Response via : Initial Calibration

